

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121222\
 Data File : VN075714.D
 Acq On : 12 Dec 2022 15:41
 Operator : JC\MD
 Sample : N6000-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GW-BR-04-138-152-120922

Quant Time: Dec 13 01:33:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120922W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 12 07:54:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	227972	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	350192	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	308285	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	129851	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	115906	47.918	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.840%
35) Dibromofluoromethane	8.171	113	103458	48.422	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.840%
50) Toluene-d8	10.571	98	409080	50.983	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.960%
62) 4-Bromofluorobenzene	12.853	95	125126	45.610	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.220%
Target Compounds						Qvalue
12) 1,1-Dichloroethene	4.359	96	16809	8.607	ug/l	91
27) cis-1,2-Dichloroethene	7.500	96	28077	10.597	ug/l	92
44) Trichloroethene	9.359	130	10897	4.436	ug/l	92
52) Toluene	10.635	92	542666	94.297	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121222\
Data File : VN075714.D
Acq On : 12 Dec 2022 15:41
Operator : JC\MD
Sample : N6000-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GW-BR-04-138-152-120922

Quant Time: Dec 13 01:33:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120922W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 12 07:54:35 2022
Response via : Initial Calibration

