

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121222\
 Data File : VN075721.D
 Acq On : 12 Dec 2022 18:29
 Operator : JC\MD
 Sample : N5978-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TOLL-MW-05

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/13/2022
 Supervised By : Mahesh Dadoda 12/13/2022

Quant Time: Dec 13 01:34:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120922W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 12 07:54:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	239794	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	367689	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	327302	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	141134	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	122672	48.215	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	96.420%		
35) Dibromofluoromethane	8.171	113	111002	49.481	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.960%		
50) Toluene-d8	10.571	98	434019	51.518	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	103.040%		
62) 4-Bromofluorobenzene	12.853	95	135836	47.158	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	94.320%		
Target Compounds					Qvalue	
16) Acetone	4.454	43	1770m	2.054	ug/l	
19) Methyl tert-butyl Ether	5.806	73	11556	1.672	ug/l #	88
52) Toluene	10.636	92	1665	0.276	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121222\
Data File : VN075721.D
Acq On : 12 Dec 2022 18:29
Operator : JC\MD
Sample : N5978-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TOLL-MW-05

Quant Time: Dec 13 01:34:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120922W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 12 07:54:35 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 12/13/2022
Supervised By :Mahesh Dadoda 12/13/2022

