

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121222\
 Data File : VN075737.D
 Acq On : 13 Dec 2022 03:41
 Operator : JC\MD
 Sample : PB149517TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149517TB

Quant Time: Dec 13 05:35:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120922W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 12 07:54:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	243900	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	383525	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	342512	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	146256	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	121823	47.075	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.160%
35) Dibromofluoromethane	8.177	113	110881	47.386	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.780%
50) Toluene-d8	10.570	98	447830	50.962	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.920%
62) 4-Bromofluorobenzene	12.853	95	141608	47.132	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.260%
Target Compounds						
					Qvalue	
16) Acetone	4.441	43	15263	17.416	ug/l	98
20) Methylene Chloride	5.283	84	10099	3.865	ug/l	93
37) Ethyl Acetate	7.571	43	21002	8.419	ug/l	97
43) Isopropyl Acetate	8.694	43	107065	25.708	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121222\
Data File : VN075737.D
Acq On : 13 Dec 2022 03:41
Operator : JC\MD
Sample : PB149517TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149517TB

Quant Time: Dec 13 05:35:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120922W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 12 07:54:35 2022
Response via : Initial Calibration

