

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121222\
 Data File : VN075740.D
 Acq On : 13 Dec 2022 04:53
 Operator : JC\MD
 Sample : PB149517ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149517ZHE#03

Quant Time: Dec 13 05:35:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120922W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 12 07:54:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	256687	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	411369	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	371217	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	164753	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	129579	47.578	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.160%
35) Dibromofluoromethane	8.171	113	121584	48.443	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.880%
50) Toluene-d8	10.571	98	480143	50.941	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.880%
62) 4-Bromofluorobenzene	12.853	95	157019	48.724	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.440%
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	18219	19.753	ug/l	99
20) Methylene Chloride	5.289	84	10671	3.881	ug/l #	96
37) Ethyl Acetate	7.571	43	22277	8.326	ug/l	97
43) Isopropyl Acetate	8.694	43	278702	62.391	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121222\
Data File : VN075740.D
Acq On : 13 Dec 2022 04:53
Operator : JC\MD
Sample : PB149517ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149517ZHE#03

Quant Time: Dec 13 05:35:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120922W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 12 07:54:35 2022
Response via : Initial Calibration

