

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121222\
 Data File : VN075742.D
 Acq On : 13 Dec 2022 05:41
 Operator : JC\MD
 Sample : PB149517ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149517ZHE#05

Quant Time: Dec 13 06:01:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120922W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 12 07:54:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	244333	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	395680	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	357590	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	155484	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	125632	48.461	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.920%
35) Dibromofluoromethane	8.171	113	117917	48.845	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.700%
50) Toluene-d8	10.571	98	462430	51.007	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.020%
62) 4-Bromofluorobenzene	12.853	95	150986	48.709	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.420%
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	18811	21.426	ug/l	100
20) Methylene Chloride	5.289	84	11502	4.395	ug/l #	84
37) Ethyl Acetate	7.571	43	22445	8.721	ug/l	98
43) Isopropyl Acetate	8.694	43	293567	68.324	ug/l #	67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121222\
Data File : VN075742.D
Acq On : 13 Dec 2022 05:41
Operator : JC\MD
Sample : PB149517ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149517ZHE#05

Quant Time: Dec 13 06:01:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120922W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 12 07:54:35 2022
Response via : Initial Calibration

