

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121222\
 Data File : VN075743.D
 Acq On : 13 Dec 2022 06:04
 Operator : JC\MD
 Sample : PB149517ZHE#06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149517ZHE#06

Quant Time: Dec 13 06:23:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120922W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 12 07:54:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	245907	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	395130	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	356959	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	151385	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	124133	47.576	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.160%
35) Dibromofluoromethane	8.171	113	117967	48.934	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.860%
50) Toluene-d8	10.571	98	459437	50.747	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.500%
62) 4-Bromofluorobenzene	12.853	95	145993	47.164	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.320%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	18405	20.830	ug/l	97
20) Methylene Chloride	5.300	84	11367	4.315	ug/l	87
37) Ethyl Acetate	7.565	43	21716	8.449	ug/l	98
43) Isopropyl Acetate	8.688	43	251944	58.719	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121222\
Data File : VN075743.D
Acq On : 13 Dec 2022 06:04
Operator : JC\MD
Sample : PB149517ZHE#06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149517ZHE#06

Quant Time: Dec 13 06:23:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120922W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 12 07:54:35 2022
Response via : Initial Calibration

