

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121222\
 Data File : VN075744.D
 Acq On : 13 Dec 2022 06:28
 Operator : JC\MD
 Sample : PB149517ZHE#07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149517ZHE#07

Quant Time: Dec 13 08:32:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120922W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 12 07:54:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	243455	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	389399	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	348095	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	148404	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	123943	47.982	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.960%
35) Dibromofluoromethane	8.171	113	113532	47.787	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.580%
50) Toluene-d8	10.571	98	454670	50.960	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.920%
62) 4-Bromofluorobenzene	12.853	95	143067	46.899	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.800%
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	11488	13.132	ug/l	96
20) Methylene Chloride	5.295	84	7695	2.951	ug/l	95
37) Ethyl Acetate	7.565	43	19556	7.721	ug/l	96
43) Isopropyl Acetate	8.694	43	293699	69.458	ug/l #	66

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121222\
Data File : VN075744.D
Acq On : 13 Dec 2022 06:28
Operator : JC\MD
Sample : PB149517ZHE#07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149517ZHE#07

Quant Time: Dec 13 08:32:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120922W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 12 07:54:35 2022
Response via : Initial Calibration

