

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121222\
 Data File : VN075746.D
 Acq On : 13 Dec 2022 07:16
 Operator : JC\MD
 Sample : N5964-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU-518-COMP-02

Quant Time: Dec 13 08:33:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120922W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 12 07:54:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	246973	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	395296	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	358224	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	154585	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	127863	48.794	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.580%
35) Dibromofluoromethane	8.171	113	116500	48.305	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.600%
50) Toluene-d8	10.571	98	463115	51.132	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.260%
62) 4-Bromofluorobenzene	12.853	95	148942	48.097	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.200%
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	11439	12.890	ug/l	96
20) Methylene Chloride	5.300	84	8751	3.308	ug/l #	71
37) Ethyl Acetate	7.571	43	19071	7.417	ug/l #	94
43) Isopropyl Acetate	8.688	43	171511	39.956	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121222\
Data File : VN075746.D
Acq On : 13 Dec 2022 07:16
Operator : JC\MD
Sample : N5964-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU-518-COMP-02

Quant Time: Dec 13 08:33:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120922W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 12 07:54:35 2022
Response via : Initial Calibration

