

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121222\
 Data File : VN075748.D
 Acq On : 13 Dec 2022 08:04
 Operator : JC\MD
 Sample : N5964-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU-518-COMP-04

Quant Time: Dec 13 08:33:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120922W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 12 07:54:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	232725	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	371876	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	341697	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	145632	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	122879	49.763	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.520%
35) Dibromofluoromethane	8.177	113	112030	49.377	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.760%
50) Toluene-d8	10.570	98	437333	51.326	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.660%
62) 4-Bromofluorobenzene	12.847	95	141591	48.602	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.200%
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	10795	12.909	ug/l	96
18) Methyl Acetate	5.041	43	2726	1.014	ug/l #	62
20) Methylene Chloride	5.294	84	8919	3.578	ug/l	90
37) Ethyl Acetate	7.577	43	18654	7.712	ug/l	98
43) Isopropyl Acetate	8.694	43	183050	45.330	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121222\
Data File : VN075748.D
Acq On : 13 Dec 2022 08:04
Operator : JC\MD
Sample : N5964-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU-518-COMP-04

Quant Time: Dec 13 08:33:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120922W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 12 07:54:35 2022
Response via : Initial Calibration

