

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121222\
 Data File : VN075753.D
 Acq On : 13 Dec 2022 10:03
 Operator : JC\MD
 Sample : N5951-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Dec 13 13:05:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120922W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 12 07:54:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	247156	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	395847	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	336726	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	144385	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	125721	47.941	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.880%
35) Dibromofluoromethane	8.171	113	117063	48.471	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.940%
50) Toluene-d8	10.571	98	448214	49.418	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.840%
62) 4-Bromofluorobenzene	12.847	95	137564	44.361	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121222\
Data File : VN075753.D
Acq On : 13 Dec 2022 10:03
Operator : JC\MD
Sample : N5951-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Dec 13 13:05:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120922W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 12 07:54:35 2022
Response via : Initial Calibration

