

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121222\
 Data File : VN075755.D
 Acq On : 13 Dec 2022 10:51
 Operator : JC\MD
 Sample : N5951-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Dec 13 13:05:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120922W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 12 07:54:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	231960	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	367900	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	321114	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	135578	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	121363	49.311	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.620%
35) Dibromofluoromethane	8.171	113	109456	48.764	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.520%
50) Toluene-d8	10.565	98	419763	49.797	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.600%
62) 4-Bromofluorobenzene	12.847	95	131689	45.692	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.380%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121222\
Data File : VN075755.D
Acq On : 13 Dec 2022 10:51
Operator : JC\MD
Sample : N5951-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 55 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Dec 13 13:05:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120922W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 12 07:54:35 2022
Response via : Initial Calibration

