

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121223\
 Data File : VN080447.D
 Acq On : 12 Dec 2023 19:02
 Operator : JC\MD
 Sample : 05672-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Dec 13 02:36:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	219011	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	399064	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	425518	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	197465	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	173559	62.046	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	124.100%
35) Dibromofluoromethane	8.165	113	147168	58.164	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	116.320%
50) Toluene-d8	10.565	98	531121	56.284	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	112.560%
62) 4-Bromofluorobenzene	12.847	95	218351	60.936	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	121.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121223\
Data File : VN080447.D
Acq On : 12 Dec 2023 19:02
Operator : JC\MD
Sample : 05672-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Dec 13 02:36:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

