

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121223\
 Data File : VN080448.D
 Acq On : 12 Dec 2023 19:25
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 13 02:36:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.639	0.612	4.2	93	0.00
3 P	Chloromethane	0.759	0.599	21.1	85	0.00
4 C	Vinyl Chloride	0.546	0.493	9.7#	87	0.00
5 T	Bromomethane	0.246	0.211	14.2	88	0.00
6 T	Chloroethane	0.351	0.325	7.4	95	0.00
7 T	Trichlorofluoromethane	1.107	1.049	5.2	94	0.00
8 T	Diethyl Ether	0.349	0.326	6.6	92	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.525	0.503	4.2	90	0.00
10 T	Methyl Iodide	0.461	0.399	13.4	78	0.00
11 T	Tert butyl alcohol	0.075	0.070	6.7	88	0.00
12 CM	1,1-Dichloroethene	0.538	0.494	8.2#	92	0.00
13 T	Acrolein	0.082	0.073	11.0	93	0.00
14 T	Allyl chloride	0.665	0.605	9.0	89	0.00
15 T	Acrylonitrile	0.213	0.211	0.9	94	0.00
16 T	Acetone	0.113	0.164	-45.1#	116	0.00
17 T	Carbon Disulfide	1.523	1.339	12.1	88	0.00
18 T	Methyl Acetate	0.729	0.723	0.8	95	0.00
19 T	Methyl tert-butyl Ether	1.793	1.741	2.9	93	0.00
20 T	Methylene Chloride	0.606	0.541	10.7	89	0.00
21 T	trans-1,2-Dichloroethene	0.583	0.534	8.4	91	0.00
22 T	Diisopropyl ether	1.435	1.364	4.9	90	0.00
23 T	Vinyl Acetate	0.824	0.789	4.2	91	0.00
24 P	1,1-Dichloroethane	1.001	0.949	5.2	91	0.00
25 T	2-Butanone	0.218	0.241	-10.6	101	0.00
26 T	2,2-Dichloropropane	1.024	0.912	10.9	86	0.00
27 T	cis-1,2-Dichloroethene	0.660	0.613	7.1	90	0.00
28 T	Bromochloromethane	0.384	0.361	6.0	90	0.00
29 T	Tetrahydrofuran	0.156	0.153	1.9	95	0.00
30 C	Chloroform	1.112	1.073	3.5#	92	0.00
31 T	Cyclohexane	0.868	0.756	12.9	90	0.00
32 T	1,1,1-Trichloroethane	1.088	1.027	5.6	90	0.00
33 S	1,2-Dichloroethane-d4	0.639	0.679	-6.3	93	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.317	0.324	-2.2	91	0.00
36 T	1,1-Dichloropropene	0.518	0.492	5.0	91	0.00
37 T	Ethyl Acetate	0.325	0.313	3.7	94	0.00
38 T	Carbon Tetrachloride	0.619	0.586	5.3	91	0.00
39 T	Methylcyclohexane	0.538	0.484	10.0	85	0.00
40 TM	Benzene	1.497	1.395	6.8	91	0.00
41 T	Methacrylonitrile	0.187	0.189	-1.1	97	0.00
42 TM	1,2-Dichloroethane	0.554	0.535	3.4	94	0.00
43 T	Isopropyl Acetate	0.584	0.565	3.3	92	0.00
44 TM	Trichloroethene	0.404	0.368	8.9	88	0.00
45 C	1,2-Dichloropropane	0.352	0.323	8.2#	89	0.00
46 T	Dibromomethane	0.252	0.238	5.6	91	0.00
47 T	Bromodichloromethane	0.541	0.528	2.4	92	0.00
48 T	Methyl methacrylate	0.348	0.332	4.6	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121223\
 Data File : VN080448.D
 Acq On : 12 Dec 2023 19:25
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 13 02:36:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.005	0.005	0.0	89	0.00
50 S	Toluene-d8	1.182	1.208	-2.2	90	0.00
51 T	4-Methyl-2-Pentanone	0.321	0.319	0.6	95	0.00
52 CM	Toluene	0.946	0.883	6.7#	90	0.00
53 T	t-1,3-Dichloropropene	0.585	0.548	6.3	88	0.00
54 T	cis-1,3-Dichloropropene	0.626	0.586	6.4	88	0.00
55 T	1,1,2-Trichloroethane	0.340	0.330	2.9	91	0.00
56 T	Ethyl methacrylate	0.377	0.366	2.9	90	0.00
57 T	1,3-Dichloropropane	0.596	0.569	4.5	92	0.00
58 T	2-Chloroethyl Vinyl ether	0.238	0.230	3.4	78	0.00
59 T	2-Hexanone	0.237	0.236	0.4	95	0.00
60 T	Dibromochloromethane	0.403	0.386	4.2	89	0.00
61 T	1,2-Dibromoethane	0.352	0.337	4.3	93	0.00
62 S	4-Bromofluorobenzene	0.449	0.444	1.1	84	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.386	0.350	9.3	87	0.00
65 PM	Chlorobenzene	1.112	1.023	8.0	89	0.00
66 T	1,1,1,2-Tetrachloroethane	0.436	0.404	7.3	88	0.00
67 C	Ethyl Benzene	2.045	1.888	7.7#	88	0.00
68 T	m/p-Xylenes	0.756	0.702	7.1	88	0.00
69 T	o-Xylene	0.739	0.681	7.8	87	0.00
70 T	Styrene	1.065	1.011	5.1	87	0.00
71 P	Bromoform	0.318	0.284	10.7	82	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
73 T	Isopropylbenzene	4.017	3.910	2.7	87	0.00
74 T	N-amyl acetate	1.173	1.141	2.7	86	0.00
75 P	1,1,2,2-Tetrachloroethane	1.076	1.014	5.8	88	0.00
76 T	1,2,3-Trichloropropane	0.997	0.958	3.9	88	0.00
77 T	Bromobenzene	0.988	0.935	5.4	84	0.00
78 T	n-propylbenzene	4.429	4.368	1.4	85	0.00
79 T	2-Chlorotoluene	2.891	2.766	4.3	84	0.00
80 T	1,3,5-Trimethylbenzene	3.459	3.356	3.0	84	0.00
81 T	trans-1,4-Dichloro-2-butene	0.380	0.353	7.1	83	0.00
82 T	4-Chlorotoluene	2.972	2.806	5.6	84	0.00
83 T	tert-Butylbenzene	2.833	2.757	2.7	84	0.00
84 T	1,2,4-Trimethylbenzene	3.441	3.342	2.9	82	0.00
85 T	sec-Butylbenzene	3.624	3.465	4.4	82	0.00
86 T	p-Isopropyltoluene	3.272	3.151	3.7	80	0.00
87 T	1,3-Dichlorobenzene	1.741	1.572	9.7	82	0.00
88 T	1,4-Dichlorobenzene	1.757	1.614	8.1	84	0.00
89 T	n-Butylbenzene	2.733	2.513	8.0	78	0.00
90 T	Hexachloroethane	0.520	0.486	6.5	83	0.00
91 T	1,2-Dichlorobenzene	1.653	1.503	9.1	88	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.226	0.221	2.2	95	0.00
93 T	1,2,4-Trichlorobenzene	0.918	0.869	5.3	87	0.00
94 T	Hexachlorobutadiene	0.523	0.446	14.7	84	0.00
95 T	Naphthalene	2.787	2.661	4.5	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121223\
Data File : VN080448.D
Acq On : 12 Dec 2023 19:25
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Dec 13 02:36:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.886	0.813	8.2	83	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6