

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121223\
 Data File : VN080437.D
 Acq On : 12 Dec 2023 14:55
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN121223

Quant Time: Dec 13 02:31:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 12/13/2023
 Supervised By : Mahesh Dadoda 12/13/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	343828	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	538436	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	476299	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	227306	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	256573	58.425	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 116.860%			
35) Dibromofluoromethane	8.165	113	199581	58.461	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 116.920%			
50) Toluene-d8	10.565	98	741678	58.253	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 116.500%#			
62) 4-Bromofluorobenzene	12.847	95	270437	55.937	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 111.880%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	207301	47.180	ug/l	96
3) Chloromethane	2.360	50	207624	44.673	ug/l	99
4) Vinyl Chloride	2.512	62	169236	45.101	ug/l	98
5) Bromomethane	2.959	94	74043	43.767	ug/l	97
6) Chloroethane	3.118	64	104913	43.511	ug/l	96
7) Trichlorofluoromethane	3.501	101	357182	46.931	ug/l	98
8) Diethyl Ether	3.954	74	111989	46.716	ug/l	90
9) 1,1,2-Trichlorotrifluo...	4.371	101	172996	47.957	ug/l	95
10) Methyl Iodide	4.589	142	143332	45.172	ug/l	90
11) Tert butyl alcohol	5.500	59	113497	219.447	ug/l	99
12) 1,1-Dichloroethene	4.342	96	169624	45.891	ug/l	90
13) Acrolein	4.171	56	126149	222.721	ug/l	97
14) Allyl chloride	5.018	41	205772	44.996	ug/l	95
15) Acrylonitrile	5.712	53	346472	237.089	ug/l	99
16) Acetone	4.418	43	213570	274.280	ug/l	97
17) Carbon Disulfide	4.712	76	462288	44.132	ug/l	100
18) Methyl Acetate	5.012	43	233101	46.525	ug/l	97
19) Methyl tert-butyl Ether	5.789	73	584371	47.387	ug/l	100
20) Methylene Chloride	5.271	84	186726	44.810	ug/l	98
21) trans-1,2-Dichloroethene	5.783	96	182946	45.647	ug/l #	79
22) Diisopropyl ether	6.671	45	463727	46.983	ug/l	91
23) Vinyl Acetate	6.600	43	1324299	233.611	ug/l	96
24) 1,1-Dichloroethane	6.559	63	322841	46.881	ug/l	99
25) 2-Butanone	7.477	43	365916	243.941	ug/l	94
26) 2,2-Dichloropropane	7.489	77	328663	46.674	ug/l	98
27) cis-1,2-Dichloroethene	7.483	96	211397	46.598	ug/l	92
28) Bromochloromethane	7.812	49	132851	50.246	ug/l	99
29) Tetrahydrofuran	7.836	42	245519	228.573	ug/l	92
30) Chloroform	7.959	83	358993	46.964	ug/l	91
31) Cyclohexane	8.259	56	257889	43.183	ug/l	94
32) 1,1,1-Trichloroethane	8.165	97	351526	46.969	ug/l	97
36) 1,1-Dichloropropene	8.371	75	264552	47.394	ug/l	97
37) Ethyl Acetate	7.553	43	156899	44.887	ug/l	98
38) Carbon Tetrachloride	8.359	117	311963	46.773	ug/l	99
39) Methylcyclohexane	9.600	83	264512	45.615	ug/l	90
40) Benzene	8.606	78	746082	46.287	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121223\
 Data File : VN080437.D
 Acq On : 12 Dec 2023 14:55
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN121223

Quant Time: Dec 13 02:31:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 12/13/2023
 Supervised By : Mahesh Dadoda 12/13/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.771	41	92516	45.910	ug/l	94
42) 1,2-Dichloroethane	8.671	62	283551	47.504	ug/l	97
43) Isopropyl Acetate	8.688	43	289119	46.002	ug/l	97
44) Trichloroethene	9.353	130	200530	46.148	ug/l	100
45) 1,2-Dichloropropane	9.618	63	173656	45.871	ug/l	98
46) Dibromomethane	9.706	93	127327	46.833	ug/l #	90
47) Bromodichloromethane	9.888	83	283836	48.713	ug/l	100
48) Methyl methacrylate	9.677	41	172393	45.965	ug/l	91
49) 1,4-Dioxane	9.688	88	45976	808.754	ug/l #	85
51) 4-Methyl-2-Pentanone	10.441	43	798726	230.756	ug/l	100
52) Toluene	10.630	92	481149	47.252	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	303815	48.192	ug/l	98
54) cis-1,3-Dichloropropene	10.312	75	322483	47.858	ug/l	95
55) 1,1,2-Trichloroethane	11.018	97	176131	48.071	ug/l	97
56) Ethyl methacrylate	10.871	69	195642	48.229	ug/l	91
57) 1,3-Dichloropropane	11.165	76	298786	46.519	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.159	63	561265	219.169	ug/l	99
59) 2-Hexanone	11.194	43	575178	225.090	ug/l	100
60) Dibromochloromethane	11.359	129	207252	47.736	ug/l	100
61) 1,2-Dibromoethane	11.471	107	176822	46.629	ug/l	99
64) Tetrachloroethene	11.106	164	174250	47.383	ug/l	93
65) Chlorobenzene	11.894	112	503381	47.509	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.959	131	194677	46.826	ug/l	94
67) Ethyl Benzene	11.965	91	919488	47.196	ug/l	97
68) m/p-Xylenes	12.071	106	686698	95.386	ug/l	90
69) o-Xylene	12.400	106	330287	46.905	ug/l	89
70) Styrene	12.412	104	491973	48.504	ug/l	94
71) Bromoform	12.582	173	137611	45.389	ug/l #	100
73) Isopropylbenzene	12.694	105	865586	47.403	ug/l	98
74) N-amyl acetate	12.494	43	244908	45.929	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.935	83	215596	44.062	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	197616m	43.579	ug/l	
77) Bromobenzene	12.982	156	206626	45.999	ug/l	96
78) n-propylbenzene	13.035	91	958208	47.594	ug/l	99
79) 2-Chlorotoluene	13.124	91	613532	46.678	ug/l	98
80) 1,3,5-Trimethylbenzene	13.176	105	735213	46.750	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.735	75	73105	42.328	ug/l	87
82) 4-Chlorotoluene	13.224	91	635416	47.032	ug/l	95
83) tert-Butylbenzene	13.441	119	597466	46.389	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	741812	47.415	ug/l	99
85) sec-Butylbenzene	13.618	105	769403	46.700	ug/l	99
86) p-Isopropyltoluene	13.729	119	705370	47.416	ug/l	99
87) 1,3-Dichlorobenzene	13.735	146	368831	46.603	ug/l	98
88) 1,4-Dichlorobenzene	13.812	146	368710	46.152	ug/l	99
89) n-Butylbenzene	14.059	91	588713	47.388	ug/l	99
90) Hexachloroethane	14.335	117	108081	45.748	ug/l	71
91) 1,2-Dichlorobenzene	14.106	146	354036	47.109	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.723	75	46995	45.662	ug/l	99
93) 1,2,4-Trichlorobenzene	15.394	180	198019	47.457	ug/l	96
94) Hexachlorobutadiene	15.506	225	105153	44.266	ug/l	98
95) Naphthalene	15.641	128	600598	47.401	ug/l	100
96) 1,2,3-Trichlorobenzene	15.841	180	186465	46.309	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121223\
Data File : VN080437.D
Acq On : 12 Dec 2023 14:55
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN121223


Manual Integrations
APPROVED

Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023

Quant Time: Dec 13 02:31:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Instrument :
MSVOA_N
ClientSampleId :
ICNVN121223




Tgt Ion:168 Resp: 343828
Ion Ratio Lower Upper
168 100
99 63.4 53.0 79.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023




#2
Dichlorodifluoromethane
Concen: 47.180 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Tgt Ion: 85 Resp: 207301
Ion Ratio Lower Upper
85 100
87 33.7 15.8 47.4





#3
Chloromethane
Concen: 44.673 ug/l
RT: 2.360 min Scan# 69
Delta R.T. 0.001 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Instrument :
MSVOA_N
ClientSampleId :
ICNVN121223




Tgt Ion: 50 Resp: 207624
Ion Ratio Lower Upper
50 100
52 32.3 26.5 39.7

Manual Integrations
APPROVED



Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023




#4
Vinyl Chloride
Concen: 45.101 ug/l
RT: 2.512 min Scan# 95
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Tgt Ion: 62 Resp: 169236
Ion Ratio Lower Upper
62 100
64 33.9 26.2 39.2





#5
Bromomethane
Concen: 43.767 ug/l
RT: 2.959 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Instrument :
MSVOA_N
ClientSampleId :
ICVVN121223



Tgt Ion: 94 Resp: 7404
Ion Ratio Lower Upper
94 100
96 94.5 73.1 109.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023



#6
Chloroethane
Concen: 43.511 ug/l
RT: 3.118 min Scan# 198
Delta R.T. -0.006 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55


Tgt Ion: 64 Resp: 104913
Ion Ratio Lower Upper
64 100
66 29.4 25.4 38.2





#7
 Trichlorofluoromethane
 Concen: 46.931 ug/l
 RT: 3.501 min Scan# 20
 Delta R.T. 0.000 min
 Lab File: VN080437.D
 Acq: 12 Dec 2023 14:55




Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN121223



Tgt Ion:101 Resp: 35718
 Ion Ratio Lower Upper
 101 100
 103 64.8 50.7 76.1




Manual Integrations
 APPROVED


Reviewed By :John Carlone 12/13/2023
 Supervised By :Mahesh Dadoda 12/13/2023



#8
 Diethyl Ether
 Concen: 46.716 ug/l
 RT: 3.954 min Scan# 340
 Delta R.T. -0.006 min
 Lab File: VN080437.D
 Acq: 12 Dec 2023 14:55


Tgt Ion: 74 Resp: 111989
 Ion Ratio Lower Upper
 74 100
 45 77.0 34.4 103.3





#9
1,1,2-Trichlorotrifluoroethane
Concen: 47.957 ug/l
RT: 4.371 min Scan# 411
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Instrument :
MSVOA_N
ClientSampleId :
ICNVN121223




Tgt Ion:101 Resp: 172990
Ion Ratio Lower Upper
101 100
85 45.2 36.6 55.0
151 78.4 57.6 86.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023




#10
Methyl Iodide
Concen: 45.172 ug/l
RT: 4.589 min Scan# 448
Delta R.T. 0.006 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Tgt Ion:142 Resp: 143332
Ion Ratio Lower Upper
142 100
127 55.1 37.6 56.4
141 15.6 11.4 17.0





#11
Tert butyl alcohol
Concen: 219.447 ug/l
RT: 5.500 min Scan# 605
Delta R.T. -0.011 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Instrument :
MSVOA_N
ClientSampleId :
ICNVN121223



Tgt Ion: 59 Resp: 11349
Ion Ratio Lower Upper
59 100
57 10.8 8.2 12.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023



#12
1,1-Dichloroethene
Concen: 45.891 ug/l
RT: 4.342 min Scan# 406
Delta R.T. 0.006 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55

Tgt Ion: 96 Resp: 169624
Ion Ratio Lower Upper
96 100
61 157.4 112.7 169.1
98 62.3 51.2 76.8





#13

Acrolein

Concen: 222.721 ug/l

RT: 4.171 min Scan# 31

Delta R.T. -0.006 min

Lab File: VN080437.D

Acq: 12 Dec 2023 14:55

Instrument :

MSVOA_N

ClientSampleId :

ICNVN121223

Tgt Ion: 56 Resp: 126149

Ion Ratio Lower Upper

56 100





55 70.5 54.5 81.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 12/13/2023

Supervised By :Mahesh Dadoda 12/13/2023



#14

Allyl chloride

Concen: 44.996 ug/l

RT: 5.018 min Scan# 521

Delta R.T. 0.000 min

Lab File: VN080437.D

Acq: 12 Dec 2023 14:55




Tgt Ion: 41 Resp: 205772


Ion Ratio Lower Upper

41 100

39 79.9 58.8 88.2


76 50.0 39.1 58.7





#15
Acrylonitrile
Concen: 237.089 ug/l
RT: 5.712 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Instrument :
MSVOA_N
ClientSampleId :
ICNVN121223




Tgt Ion: 53 Resp: 346472
Ion Ratio Lower Upper
53 100
52 80.8 65.3 97.9
51 35.6 28.9 43.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023




#16
Acetone
Concen: 274.280 ug/l
RT: 4.418 min Scan# 419
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Tgt Ion: 43 Resp: 213570
Ion Ratio Lower Upper
43 100
58 33.3 28.0 42.0





#17
Carbon Disulfide
Concen: 44.132 ug/l
RT: 4.712 min Scan# 469
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Instrument :
MSVOA_N
Client Sample Id :
ICV121223




Tgt Ion: 76 Resp: 46228
Ion Ratio Lower Upper
76 100
78 9.2 7.4 11.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023




#18
Methyl Acetate
Concen: 46.525 ug/l
RT: 5.012 min Scan# 520
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Tgt Ion: 43 Resp: 233101
Ion Ratio Lower Upper
43 100
74 30.9 26.2 39.4





#19
Methyl tert-butyl Ether
Concen: 47.387 ug/l
RT: 5.789 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Instrument :
MSVOA_N
ClientSampleId :
ICNVN121223



Tgt Ion: 73 Resp: 58437
Ion Ratio Lower Upper
73 100
57 19.4 15.6 23.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023



#20
Methylene Chloride
Concen: 44.810 ug/l
RT: 5.271 min Scan# 564
Delta R.T. -0.006 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55


Tgt Ion: 84 Resp: 186726
Ion Ratio Lower Upper
84 100
49 102.7 79.7 119.5
51 30.8 25.2 37.8
86 65.8 53.3 79.9





#21
trans-1,2-Dichloroethene
Concen: 45.647 ug/l
RT: 5.783 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Instrument :
MSVOA_N
ClientSampleId :
ICNVN121223




Tgt Ion: 96 Resp: 182940
Ion Ratio Lower Upper
96 100
61 140.5 90.1 135.1
98 66.0 46.4 69.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023



#22
Diisopropyl ether
Concen: 46.983 ug/l
RT: 6.671 min Scan# 802
Delta R.T. 0.006 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Tgt Ion: 45 Resp: 463727
Ion Ratio Lower Upper
45 100
43 50.4 45.5 68.3
87 33.4 31.8 47.8
59 15.7 11.6 17.4





#23
 Vinyl Acetate
 Concen: 233.611 ug/l
 RT: 6.600 min Scan# 790
 Delta R.T. 0.000 min
 Lab File: VN080437.D
 Acq: 12 Dec 2023 14:55




Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN121223




Tgt Ion: 43 Resp: 1324299
 Ion Ratio Lower Upper
 43 100
 86 14.4 13.0 19.6

Manual Integrations
 APPROVED



Reviewed By :John Carlone 12/13/2023
 Supervised By :Mahesh Dadoda 12/13/2023




#24
 1,1-Dichloroethane
 Concen: 46.881 ug/l
 RT: 6.559 min Scan# 783
 Delta R.T. -0.006 min
 Lab File: VN080437.D
 Acq: 12 Dec 2023 14:55




Tgt Ion: 63 Resp: 322841
 Ion Ratio Lower Upper
 63 100
 98 8.3 3.9 11.7
 100 4.8 2.5 7.4





#25
2-Butanone
Concen: 243.941 ug/l
RT: 7.477 min Scan# 939
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Instrument :
MSVOA_N
ClientSampleId :
ICNVN121223



Tgt Ion: 43 Resp: 365910
Ion Ratio Lower Upper
43 100
72 34.3 30.2 45.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023



#26
2,2-Dichloropropane
Concen: 46.674 ug/l
RT: 7.489 min Scan# 941
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55


Tgt Ion: 77 Resp: 328663
Ion Ratio Lower Upper
77 100
97 21.5 11.2 33.6





#27
 cis-1,2-Dichloroethene
 Concen: 46.598 ug/l
 RT: 7.483 min Scan# 940
 Delta R.T. 0.000 min
 Lab File: VN080437.D
 Acq: 12 Dec 2023 14:55




Instrument :
 MSVOA_N
 ClientSampleId :
 ICVN121223




Tgt Ion: 96 Resp: 21139
 Ion Ratio Lower Upper
 96 100
 61 142.0 0.0 255.2
 98 64.3 0.0 128.8

Manual Integrations
 APPROVED



Reviewed By :John Carlone 12/13/2023
 Supervised By :Mahesh Dadoda 12/13/2023




#28
 Bromochloromethane
 Concen: 50.246 ug/l
 RT: 7.812 min Scan# 996
 Delta R.T. 0.000 min
 Lab File: VN080437.D
 Acq: 12 Dec 2023 14:55



Tgt Ion: 49 Resp: 132851
 Ion Ratio Lower Upper
 49 100
 129 1.4 0.0 4.8
 130 93.0 75.3 112.9





#29

Tetrahydrofuran

Concen: 228.573 ug/l

RT: 7.836 min Scan# 1000

Delta R.T. 0.000 min

Lab File: VN080437.D


Acq: 12 Dec 2023 14:55

Instrument :

MSVOA_N

ClientSampleId :

ICNVN121223






Tgt Ion:	42	Resp:	245519
Ion Ratio	100	Lower	Upper
42	100		
72	56.5	49.8	74.6
71	53.0	47.4	71.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 12/13/2023

Supervised By :Mahesh Dadoda 12/13/2023



#30

Chloroform

Concen: 46.964 ug/l




RT: 7.959 min Scan# 1021


Delta R.T. -0.006 min

Lab File: VN080437.D

Acq: 12 Dec 2023 14:55


Tgt Ion:	83	Resp:	358993
Ion Ratio	100	Lower	Upper
83	100		
85	62.0	55.4	83.0





#31
Cyclohexane
Concen: 43.183 ug/l
RT: 8.259 min Scan# 1071
Delta R.T. 0.006 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Instrument :
MSVOA_N
ClientSampleId :
ICNVN121223




Tgt Ion: 56 Resp: 257889
Ion Ratio Lower Upper
56 100
69 34.8 26.8 40.2
84 95.4 81.8 122.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023




#32
1,1,1-Trichloroethane
Concen: 46.969 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55



Tgt Ion: 97 Resp: 351526
Ion Ratio Lower Upper
97 100
99 63.4 50.3 75.5
61 39.6 34.7 52.1





#33

1,2-Dichloroethane-d4

Concen: 58.425 ug/l

RT: 8.577 min Scan# 1126

Delta R.T. 0.000 min

Lab File: VN080437.D

Acq: 12 Dec 2023 14:55

Instrument :

MSVOA_N

ClientSampleId :

ICNVN121223

Tgt Ion: 65 Resp: 25657

Ion Ratio Lower Upper

65 100





67 49.3 0.0 106.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 12/13/2023

Supervised By :Mahesh Dadoda 12/13/2023



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. 0.000 min

Lab File: VN080437.D

Acq: 12 Dec 2023 14:55




Tgt Ion:114 Resp: 538436


Ion Ratio Lower Upper

114 100

63 19.8 0.0 36.4

88 17.5 0.0 30.6





#35

Dibromofluoromethane

Concen: 58.461 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.000 min

Lab File: VN080437.D

Acq: 12 Dec 2023 14:55

Instrument :

MSVOA_N

ClientSampleId :

ICNVN121223

Tgt Ion: 113 Resp: 19958





Ion	Ratio	Lower	Upper
113	100		
111	100.6	82.6	123.8
192	19.0	12.7	19.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 12/13/2023

Supervised By :Mahesh Dadoda 12/13/2023



#36

1,1-Dichloropropene

Concen: 47.394 ug/l

RT: 8.371 min Scan# 1091




Delta R.T. 0.000 min


Lab File: VN080437.D

Acq: 12 Dec 2023 14:55

Tgt Ion: 75 Resp: 264552


Ion	Ratio	Lower	Upper
75	100		
110	35.0	19.1	57.3
77	31.2	25.0	37.4





#37
Ethyl Acetate
Concen: 44.887 ug/l
RT: 7.553 min Scan# 952
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Instrument :
MSVOA_N
ClientSampleId :
ICNVN121223



Tgt Ion: 43 Resp: 156899
Ion Ratio Lower Upper
43 100
61 15.7 13.2 19.8
70 14.7 11.3 16.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023



#38
Carbon Tetrachloride
Concen: 46.773 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55

Tgt Ion:117 Resp: 311963
Ion Ratio Lower Upper
117 100
119 96.7 77.3 115.9
121 32.0 23.3 34.9





#39

Methylcyclohexane

Concen: 45.615 ug/l

RT: 9.600 min Scan# 11300

Delta R.T. 0.000 min

Lab File: VN080437.D

Acq: 12 Dec 2023 14:55

Instrument :

MSVOA_N

ClientSampleId :

ICVVN121223





Tgt Ion:	83	Resp:	26451
Ion Ratio	Lower	Upper	
83	100		
55	72.7	49.1	73.7
98	47.6	40.6	61.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 12/13/2023

Supervised By :Mahesh Dadoda 12/13/2023



#40

Benzene

Concen: 46.287 ug/l




RT: 8.606 min Scan# 1131


Delta R.T. 0.000 min

Lab File: VN080437.D

Acq: 12 Dec 2023 14:55


Tgt Ion:	78	Resp:	746082
Ion Ratio	Lower	Upper	
78	100		
77	23.9	19.0	28.6





#41
Methacrylonitrile
Concen: 45.910 ug/l
RT: 7.771 min Scan# 989
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Instrument :
MSVOA_N
ClientSampleId :
ICNVN121223




Tgt Ion: 41 Resp: 92510
Ion Ratio Lower Upper
41 100
39 55.9 42.6 64.0
67 88.6 77.5 116.3
52 35.1 28.9 43.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023




#42
1,2-Dichloroethane
Concen: 47.504 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Tgt Ion: 62 Resp: 283551
Ion Ratio Lower Upper
62 100
98 8.8 0.0 19.8





#43
Isopropyl Acetate
Concen: 46.002 ug/l
RT: 8.688 min Scan# 1145
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Instrument :
MSVOA_N
ClientSampleId :
ICNVN121223




Tgt Ion: 43 Resp: 289119
Ion Ratio Lower Upper
43 100
61 32.9 24.7 37.1
87 17.2 14.6 21.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023




#44
Trichloroethene
Concen: 46.148 ug/l
RT: 9.353 min Scan# 1258
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55



Tgt Ion:130 Resp: 200530
Ion Ratio Lower Upper
130 100
95 102.8 0.0 204.6





#45

1,2-Dichloropropane

Concen: 45.871 ug/l

RT: 9.618 min Scan# 1303

Delta R.T. -0.006 min

Lab File: VN080437.D

Acq: 12 Dec 2023 14:55

Instrument :

MSVOA_N

ClientSampleId :

ICNVN121223

Tgt Ion: 63 Resp: 173650

Ion Ratio Lower Upper

63 100





65 31.2 23.9 35.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 12/13/2023

Supervised By :Mahesh Dadoda 12/13/2023



#46

Dibromomethane

Concen: 46.833 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.006 min

Lab File: VN080437.D

Acq: 12 Dec 2023 14:55




Tgt Ion: 93 Resp: 127327


Ion Ratio Lower Upper

93 100

95 80.1 63.8 95.8


174 99.2 65.6 98.4#





#47
Bromodichloromethane
Concen: 48.713 ug/l
RT: 9.888 min Scan# 1349
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Instrument :
MSVOA_N
ClientSampleId :
ICNVN121223




Tgt Ion: 83 Resp: 283830
Ion Ratio Lower Upper
83 100
85 61.8 49.6 74.4
127 8.6 7.0 10.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023




#48
Methyl methacrylate
Concen: 45.965 ug/l
RT: 9.677 min Scan# 1313
Delta R.T. -0.006 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55



Tgt Ion: 41 Resp: 172393
Ion Ratio Lower Upper
41 100
69 103.1 90.4 135.6
39 62.1 54.6 81.8





#49

1,4-Dioxane

Concen: 808.754 ug/l

RT: 9.688 min Scan# 1315

Delta R.T. 0.000 min

Lab File: VN080437.D

Acq: 12 Dec 2023 14:55

Instrument :

MSVOA_N

ClientSampleId :

ICNVN121223





Tgt Ion:	88	Resp:	45970
Ion	Ratio	Lower	Upper
88	100		
43	30.7	17.0	25.6
58	68.6	47.0	70.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 12/13/2023

Supervised By :Mahesh Dadoda 12/13/2023



#50

Toluene-d8

Concen: 58.253 ug/l




RT: 10.565 min Scan# 1464


Delta R.T. 0.000 min

Lab File: VN080437.D

Acq: 12 Dec 2023 14:55


Tgt Ion:	98	Resp:	741678
Ion	Ratio	Lower	Upper
98	100		
100	65.9	52.0	78.0





#51
4-Methyl-2-Pentanone
Concen: 230.756 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Instrument :
MSVOA_N
ClientSampleId :
ICNVN121223




Tgt Ion: 43 Resp: 798720
Ion Ratio Lower Upper
43 100
58 45.8 36.6 54.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023




#52
Toluene
Concen: 47.252 ug/l
RT: 10.630 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55



Tgt Ion: 92 Resp: 481149
Ion Ratio Lower Upper
92 100
91 170.1 137.4 206.0





#53
t-1,3-Dichloropropene
Concen: 48.192 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55

Instrument :

MSVOA_N

ClientSampleId :





ICNVN121223

Tgt Ion: 75 Resp: 303815
Ion Ratio Lower Upper
75 100
77 32.6 25.4 38.0

Manual Integrations




APPROVED


Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023



#54
cis-1,3-Dichloropropene
Concen: 47.858 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55


Tgt Ion: 75 Resp: 322483
Ion Ratio Lower Upper
75 100
77 30.5 26.1 39.1
39 42.7 31.6 47.4





#55
1,1,2-Trichloroethane
Concen: 48.071 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Instrument :
MSVOA_N
ClientSampleId :
ICNVN121223



Tgt Ion:	97	Resp:	17613
Ion	Ratio	Lower	Upper
97	100		
83	83.6	65.1	97.7
85	56.9	43.4	65.2
99	63.3	49.2	73.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023



#56
Ethyl methacrylate
Concen: 48.229 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55

Tgt Ion:	69	Resp:	195642
Ion	Ratio	Lower	Upper
69	100		
41	54.8	39.6	59.4
39	29.0	28.6	43.0





#57

1,3-Dichloropropane

Concen: 46.519 ug/l

RT: 11.165 min Scan# 1566

Delta R.T. 0.006 min

Lab File: VN080437.D

Acq: 12 Dec 2023 14:55

Instrument :

MSVOA_N

ClientSampleId :

ICNVN121223

Tgt Ion: 76 Resp: 298780

Ion Ratio Lower Upper

76 100





78 33.4 26.1 39.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 12/13/2023

Supervised By :Mahesh Dadoda 12/13/2023



#58

2-Chloroethyl Vinyl ether

Concen: 219.169 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. 0.000 min

Lab File: VN080437.D




Acq: 12 Dec 2023 14:55


Tgt Ion: 63 Resp: 561265

Ion Ratio Lower Upper

63 100


106 32.1 26.2 39.4





#59
2-Hexanone
Concen: 225.090 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Instrument :
MSVOA_N
ClientSampleId :
ICNVN121223



Tgt Ion: 43 Resp: 575178
Ion Ratio Lower Upper
43 100
58 63.0 31.7 95.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023



#60
Dibromochloromethane
Concen: 47.736 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55


Tgt Ion:129 Resp: 207252
Ion Ratio Lower Upper
129 100
127 77.1 38.6 116.0





#61
1,2-Dibromoethane
Concen: 46.629 ug/l
RT: 11.471 min Scan# 1618
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Instrument :
MSVOA_N
ClientSampleId :
ICVVN121223




Tgt Ion:107 Resp: 17682
Ion Ratio Lower Upper
107 100
109 94.3 76.6 114.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023




#62
4-Bromofluorobenzene
Concen: 55.937 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Tgt Ion: 95 Resp: 270437
Ion Ratio Lower Upper
95 100
174 75.0 0.0 143.8
176 73.8 0.0 135.8





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Instrument :
MSVOA_N
ClientSampleId :
ICNVN121223



Tgt Ion:117 Resp: 476299
Ion Ratio Lower Upper
117 100
82 60.2 44.0 66.0
119 32.9 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023



#64
Tetrachloroethene
Concen: 47.383 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. 0.006 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55


Tgt Ion:164 Resp: 174250
Ion Ratio Lower Upper
164 100
166 121.0 101.4 152.0
129 93.9 82.4 123.6
131 91.3 78.2 117.4





#65
Chlorobenzene
Concen: 47.509 ug/l
RT: 11.894 min Scan# 1690
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Instrument :
MSVOA_N
ClientSampleId :
ICNVN121223




Tgt Ion:112 Resp: 50338
Ion Ratio Lower Upper
112 100
114 31.8 25.0 37.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023




#66
1,1,1,2-Tetrachloroethane
Concen: 46.826 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55



Tgt Ion:131 Resp: 194677
Ion Ratio Lower Upper
131 100
133 96.0 46.4 139.1
119 59.6 34.1 102.2









#67
Ethyl Benzene
Concen: 47.196 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55

Instrument :
MSVOA_N
ClientSampleId :
ICNVN121223

Tgt Ion: 91 Resp: 919488
Ion Ratio Lower Upper
91 100
106 30.5 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023



#68
m/p-Xylenes
Concen: 95.386 ug/l
RT: 12.071 min Scan# 1720
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55


Tgt Ion:106 Resp: 686698
Ion Ratio Lower Upper
106 100
91 215.2 159.6 239.4





#69
o-Xylene
Concen: 46.905 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Instrument :
MSVOA_N
ClientSampleId :
ICNVN121223



Tgt Ion:106 Resp: 33028
Ion Ratio Lower Upper
106 100
91 227.5 105.2 315.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023



#70
Styrene
Concen: 48.504 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55


Tgt Ion:104 Resp: 491973
Ion Ratio Lower Upper
104 100
78 56.6 39.5 59.3
103 55.7 43.2 64.8





#71
Bromoform
Concen: 45.389 ug/l
RT: 12.582 min Scan# 1806
Delta R.T. 0.006 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Instrument :
MSVOA_N
ClientSampleId :
ICVNV121223




Tgt Ion:173 Resp: 137611
Ion Ratio Lower Upper
173 100
175 48.8 24.3 73.0
254 0.1 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55



Tgt Ion:152 Resp: 227306
Ion Ratio Lower Upper
152 100
115 65.5 32.3 96.8
150 173.6 0.0 355.8





#73

Isopropylbenzene

Concen: 47.403 ug/l

RT: 12.694 min Scan# 1826

Delta R.T. 0.000 min

Lab File: VN080437.D

Acq: 12 Dec 2023 14:55

Instrument :

MSVOA_N

ClientSampleId :

ICNVN121223

Tgt Ion:105 Resp: 865580

Ion Ratio Lower Upper

105 100





120 25.0 13.1 39.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 12/13/2023

Supervised By :Mahesh Dadoda 12/13/2023



#74

N-ethyl acetate

Concen: 45.929 ug/l

RT: 12.494 min Scan# 1792

Delta R.T. 0.000 min

Lab File: VN080437.D

Acq: 12 Dec 2023 14:55

Tgt Ion: 43 Resp: 244908




Ion Ratio Lower Upper


43 100

70 53.7 43.4 65.2

55 28.5 21.4 32.0

61 28.5 22.8 34.2









#75
1,1,2,2-Tetrachloroethane
Concen: 44.062 ug/l
RT: 12.935 min Scan# 1867
Delta R.T. -0.006 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55

Instrument :
MSVOA_N
ClientSampleId :
ICNVN121223

Tgt Ion: 83 Resp: 215590
Ion Ratio Lower Upper
83 100
131 11.5 5.0 15.0
85 64.5 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023



#76
1,2,3-Trichloropropane
Concen: 43.579 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55

Tgt Ion: 75 Resp: 197616
Ion Ratio Lower Upper
75 100
77 216.8 84.5 253.5





#77

Bromobenzene

Concen: 45.999 ug/l

RT: 12.982 min Scan# 1875

Delta R.T. 0.000 min

Lab File: VN080437.D

Acq: 12 Dec 2023 14:55

Instrument :

MSVOA_N

ClientSampleId :

ICNVN121223





Tgt Ion:156 Resp: 206620
 Ion Ratio Lower Upper
 156 100
 77 207.4 107.7 323.3
 158 98.3 48.4 145.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 12/13/2023

Supervised By :Mahesh Dadoda 12/13/2023



#78

n-propylbenzene

Concen: 47.594 ug/l




RT: 13.035 min Scan# 1884


Delta R.T. 0.000 min

Lab File: VN080437.D

Acq: 12 Dec 2023 14:55


Tgt Ion: 91 Resp: 958208
 Ion Ratio Lower Upper
 91 100
 120 22.5 11.5 34.4





#79
2-Chlorotoluene
Concen: 46.678 ug/l
RT: 13.124 min Scan# 1899
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Instrument :
MSVOA_N
ClientSampleId :
ICNVN121223




Tgt Ion: 91 Resp: 61353
Ion Ratio Lower Upper
91 100
126 31.6 16.3 48.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023




#80
1,3,5-Trimethylbenzene
Concen: 46.750 ug/l
RT: 13.176 min Scan# 1908
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Tgt Ion:105 Resp: 735213
Ion Ratio Lower Upper
105 100
120 46.8 24.1 72.3





#81
trans-1,4-Dichloro-2-butene
Concen: 42.328 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Instrument :
MSVOA_N
ClientSampleId :
ICNVN121223




Tgt Ion: 75 Resp: 7310
Ion Ratio Lower Upper
75 100
53 98.6 67.4 101.2
89 43.7 38.8 58.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023




#82
4-Chlorotoluene
Concen: 47.032 ug/l
RT: 13.224 min Scan# 1916
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Tgt Ion: 91 Resp: 635416
Ion Ratio Lower Upper
91 100
126 30.7 16.7 50.0





#83
tert-Butylbenzene
Concen: 46.389 ug/l
RT: 13.441 min Scan# 1953
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Instrument :
MSVOA_N
ClientSampleId :
ICNVN121223




Tgt Ion:119 Resp: 597460
Ion Ratio Lower Upper
119 100
91 69.7 33.9 101.7
134 25.9 13.2 39.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023




#84
1,2,4-Trimethylbenzene
Concen: 47.415 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55



Tgt Ion:105 Resp: 741812
Ion Ratio Lower Upper
105 100
120 43.2 22.0 66.0









#85
 sec-Butylbenzene
 Concen: 46.700 ug/l
 RT: 13.618 min Scan# 1983
 Delta R.T. 0.000 min
 Lab File: VN080437.D
 Acq: 12 Dec 2023 14:55

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVN121223

Tgt Ion:105 Resp: 76940
 Ion Ratio Lower Upper
 105 100
 134 20.1 9.8 29.3




Manual Integrations
 APPROVED


Reviewed By :John Carlone 12/13/2023
 Supervised By :Mahesh Dadoda 12/13/2023



#86
 p-Isopropyltoluene
 Concen: 47.416 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.000 min
 Lab File: VN080437.D
 Acq: 12 Dec 2023 14:55


Tgt Ion:119 Resp: 705370
 Ion Ratio Lower Upper
 119 100
 134 25.2 12.6 37.8
 91 26.2 12.4 37.2





#87
 1,3-Dichlorobenzene
 Concen: 46.603 ug/l
 RT: 13.735 min Scan# 2003
 Delta R.T. 0.000 min
 Lab File: VN080437.D
 Acq: 12 Dec 2023 14:55




Instrument :
 MSVOA_N
 ClientSampleId :
 ICVN121223




Tgt Ion:146 Resp: 36883
 Ion Ratio Lower Upper
 146 100
 111 44.7 22.1 66.3
 148 62.2 32.0 96.0

Manual Integrations
 APPROVED



Reviewed By :John Carlone 12/13/2023
 Supervised By :Mahesh Dadoda 12/13/2023




#88
 1,4-Dichlorobenzene
 Concen: 46.152 ug/l
 RT: 13.812 min Scan# 2016
 Delta R.T. 0.000 min
 Lab File: VN080437.D
 Acq: 12 Dec 2023 14:55



Tgt Ion:146 Resp: 368710
 Ion Ratio Lower Upper
 146 100
 111 42.6 20.6 61.8
 148 63.7 31.6 94.8





#89

n-Butylbenzene

Concen: 47.388 ug/l

RT: 14.059 min Scan# 2058

Delta R.T. 0.000 min

Lab File: VN080437.D

Acq: 12 Dec 2023 14:55

Instrument :

MSVOA_N

ClientSampleId :

ICNVN121223





Tgt Ion:	91	Resp:	58871
Ion Ratio	Lower	Upper	
91	100		
92	53.8	26.4	79.0
134	25.5	12.6	37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 12/13/2023

Supervised By :Mahesh Dadoda 12/13/2023



#90

Hexachloroethane

Concen: 45.748 ug/l




RT: 14.335 min Scan# 2105


Delta R.T. 0.000 min

Lab File: VN080437.D

Acq: 12 Dec 2023 14:55


Tgt Ion:	117	Resp:	108081
Ion Ratio	Lower	Upper	
117	100		
201	80.5	29.3	88.0





#91
1,2-Dichlorobenzene
Concen: 47.109 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Instrument :
MSVOA_N
ClientSampleId :
ICNVN121223



Tgt Ion:146 Resp: 354030
Ion Ratio Lower Upper
146 100
111 45.7 22.0 66.0
148 63.4 31.8 95.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023



#92
1,2-Dibromo-3-Chloropropane
Concen: 45.662 ug/l
RT: 14.723 min Scan# 2171
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55


Tgt Ion: 75 Resp: 46995
Ion Ratio Lower Upper
75 100
155 78.8 39.1 117.2
157 99.6 50.3 150.9





#93
 1,2,4-Trichlorobenzene
 Concen: 47.457 ug/l
 RT: 15.394 min Scan# 2285
 Delta R.T. 0.000 min
 Lab File: VN080437.D
 Acq: 12 Dec 2023 14:55




Instrument :
 MSVOA_N
 ClientSampleId :
 ICVN121223



Tgt Ion:180 Resp: 198019
 Ion Ratio Lower Upper
 180 100
 182 96.3 46.1 138.2
 145 34.6 16.6 49.8




Manual Integrations
 APPROVED


Reviewed By :John Carlone 12/13/2023
 Supervised By :Mahesh Dadoda 12/13/2023



#94
 Hexachlorobutadiene
 Concen: 44.266 ug/l
 RT: 15.506 min Scan# 2304
 Delta R.T. 0.006 min
 Lab File: VN080437.D
 Acq: 12 Dec 2023 14:55


Tgt Ion:225 Resp: 105153
 Ion Ratio Lower Upper
 225 100
 223 62.7 31.3 93.8
 227 61.6 31.9 95.9





#95
Naphthalene
Concen: 47.401 ug/l
RT: 15.641 min Scan# 2327
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55




Instrument :
MSVOA_N
ClientSampleId :
ICNVN121223




Tgt Ion:128 Resp: 600598
Ion Ratio Lower Upper
128 100
127 12.6 10.0 15.0
129 11.0 8.7 13.1

Manual Integrations
APPROVED

Reviewed By :John Carlone 12/13/2023
Supervised By :Mahesh Dadoda 12/13/2023



#96
1,2,3-Trichlorobenzene
Concen: 46.309 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. 0.000 min
Lab File: VN080437.D
Acq: 12 Dec 2023 14:55



Tgt Ion:180 Resp: 186465
Ion Ratio Lower Upper
180 100
182 96.1 46.2 138.6
145 36.0 17.5 52.6

