

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121224\
 Data File : VN085182.D
 Acq On : 12 Dec 2024 11:11
 Operator : JC\MD
 Sample : VN1212WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1212WBL01

Quant Time: Dec 13 04:21:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 23 02:54:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	148964	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	257482	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	222289	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	90152	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	109819	51.755	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.520%
35) Dibromofluoromethane	8.165	113	90403	52.064	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.120%
50) Toluene-d8	10.565	98	302075	48.544	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.080%
62) 4-Bromofluorobenzene	12.847	95	101700	43.702	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	87.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121224\
Data File : VN085182.D
Acq On : 12 Dec 2024 11:11
Operator : JC\MD
Sample : VN1212WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1212WBL01

Quant Time: Dec 13 04:21:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 23 02:54:10 2024
Response via : Initial Calibration

