

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121224\
 Data File : VN085186.D
 Acq On : 12 Dec 2024 12:58
 Operator : JC\MD
 Sample : P5239-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-1

Quant Time: Dec 13 04:23:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 23 02:54:10 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	144040	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	254219	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	241017	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	95191	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	104222	50.797	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	101.600%	
35) Dibromofluoromethane	8.165	113	85502	49.874	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	99.740%	
50) Toluene-d8	10.565	98	313883	51.089	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	102.180%	
62) 4-Bromofluorobenzene	12.847	95	109576	47.691	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	95.380%	
Target Compounds						
						Qvalue
16) Acetone	4.430	43	16716	27.974	ug/l	93
20) Methylene Chloride	5.271	84	2167	1.188	ug/l #	73
43) Isopropyl Acetate	8.688	43	104865	23.915	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121224\
Data File : VN085186.D
Acq On : 12 Dec 2024 12:58
Operator : JC\MD
Sample : P5239-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-1

Quant Time: Dec 13 04:23:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 23 02:54:10 2024
Response via : Initial Calibration

