

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121318\
 Data File : VN052867.D
 Acq On : 13 Dec 2018 16:28
 Operator : MD\SY
 Sample : J6319-05DL 10X
 Misc : 5.67g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 17 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 059_BT-3-1DL

Quant Time: Dec 14 09:07:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 13 01:08:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1288276	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1955096	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1751561	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	797247	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	634728	46.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.68%	
35) Dibromofluoromethane	7.59	113	542498	49.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.42%	
50) Toluene-d8	10.09	98	2509797	53.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.08%	
62) 4-Bromofluorobenzene	12.40	95	862879	53.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.64%	
Target Compounds						
31) Cyclohexane	7.65	56	398224	16.770	ug/l	93
39) Methylcyclohexane	9.08	83	1312219	59.393	ug/l	99
73) Isopropylbenzene	12.25	105	176864	2.971	ug/l	100
78) n-propylbenzene	12.59	91	487563	7.182	ug/l	99
85) sec-Butylbenzene	13.17	105	99428	1.732	ug/l #	50
89) n-Butylbenzene	13.62	91	165942	3.963	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121318\
Data File : VN052867.D
Acq On : 13 Dec 2018 16:28
Operator : MD\SY
Sample : J6319-05DL 10X
Misc : 5.67g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 17 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
059_BT-3-1DL

Quant Time: Dec 14 09:07:13 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 13 01:08:35 2018
Response via : Initial Calibration

