

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121318\
 Data File : VN052873.D
 Acq On : 13 Dec 2018 19:08
 Operator : MD\SY
 Sample : PB115614ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115614ZHE#04

Quant Time: Dec 14 03:19:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 13 01:08:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1256491	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1942553	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1760296	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	748964	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	640934	47.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.94%	
35) Dibromofluoromethane	7.59	113	558206	51.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.96%	
50) Toluene-d8	10.09	98	2369212	50.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.78%	
62) 4-Bromofluorobenzene	12.40	95	799265	50.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.34%	
Target Compounds						
16) Acetone	3.82	43	18542	6.179	ug/l	94
20) Methylene Chloride	4.55	84	28159	2.085	ug/l	95
43) Isopropyl Acetate	8.17	43	76018	3.450	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121318\
Data File : VN052873.D
Acq On : 13 Dec 2018 19:08
Operator : MD\SY
Sample : PB115614ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115614ZHE#04

Quant Time: Dec 14 03:19:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 13 01:08:35 2018
Response via : Initial Calibration

