

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121318\
 Data File : VN052874.D
 Acq On : 13 Dec 2018 19:34
 Operator : MD\SY
 Sample : J6373-10DL 100X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 OILY-WATERDL

Quant Time: Dec 14 05:31:46 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 13 01:08:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1261098	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1943072	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1724887	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	715796	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	629518	46.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.90%	
35) Dibromofluoromethane	7.59	113	556658	51.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.64%	
50) Toluene-d8	10.09	98	2358790	50.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.32%	
62) 4-Bromofluorobenzene	12.40	95	778966	48.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.76%	
Target Compounds						
16) Acetone	3.82	43	1038196	344.734	ug/l	98
52) Toluene	10.16	92	536369	16.490	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121318\
Data File : VN052874.D
Acq On : 13 Dec 2018 19:34
Operator : MD\SY
Sample : J6373-10DL 100X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
OILY-WATERDL

Quant Time: Dec 14 05:31:46 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 13 01:08:35 2018
Response via : Initial Calibration

