

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121318\
 Data File : VN052871.D
 Acq On : 13 Dec 2018 18:14
 Operator : MD\SY
 Sample : PB115614ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115614ZHE#02

Quant Time: Dec 14 03:15:36 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 13 01:08:35 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1260440	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1923293	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1711322	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	683753	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	629711	46.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.98%	
35) Dibromofluoromethane	7.59	113	551531	51.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.74%	
50) Toluene-d8	10.09	98	2350133	50.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.98%	
62) 4-Bromofluorobenzene	12.40	95	769435	48.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.56%	
Target Compounds						
16) Acetone	3.82	43	17771	5.904	ug/l	99
20) Methylene Chloride	4.56	84	28328	2.091	ug/l	99
43) Isopropyl Acetate	8.17	43	66343	3.041	ug/l #	80


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121318\
Data File : VN052871.D
Acq On : 13 Dec 2018 18:14
Operator : MD\SY
Sample : PB115614ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115614ZHE#02

Quant Time: Dec 14 03:15:36 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 13 01:08:35 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. -0.00 min

Lab File: VN052871.D

Acq: 13 Dec 2018 18:14

Instrument :

MSVOA_N

ClientSampleId :


PB115614ZHE#02

Tot Ion:168 Resp: 1260440

Ion Ratio Lower Upper

168 100

99 50.6 40.2 60.2



Abundance Ion 167.90 (167.60 to 168.60): VN0

Ion 98.90 (98.60 to 99.60): VN0

7.67

600000

500000

400000



300000

200000

100000

0

Time-->



#16

Acetone

Concen: 5.904 ug/l

RT: 3.82 min Scan# 637

Delta R.T. 0.01 min

Lab File: VN052871.D


Acq: 13 Dec 2018 18:14

Tgt Ion: 43 Resp: 17771

Ion Ratio Lower Upper

43 100

58 32.0 25.4 38.0



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0

3.82


6000


4000

2000

0

Time-->









#20
Methylene Chloride
Concen: 2.091 ug/l
RT: 4.56 min Scan# 866
Delta R.T. 0.00 min
Lab File: VN052871.D
Acq: 13 Dec 2018 18:14




Instrument :
MSVOA_N
ClientSampleId :
PB115614ZHE#02


Tot Ion: 84 Resp: 28328
Ion Ratio Lower Upper
84 100
49 122.0 96.8 145.2
51 35.8 29.9 44.9
86 63.4 51.8 77.6



#33
1,2-Dichloroethane-d4
Concen: Below ug/l
RT: 8.03 min Scan# 1945
Delta R.T. -0.00 min
Lab File: VN052871.D
Acq: 13 Dec 2018 18:14

Tgt Ion: 65 Resp: 629711
Ion Ratio Lower Upper
65 100
67 55.2 0.0 110.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.59 min Scan# 2119

Delta R.T. -0.00 min

Lab File: VN052871.D

Acq: 13 Dec 2018 18:14

Instrument :

MSVOA_N

ClientSampleId :

PB115614ZHE#02


Tot Ion:114 Resp: 1923293

Ion Ratio Lower Upper

114 100

63 20.1 0.0 39.8




88 15.4 0.0 31.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN

Ion 87.90 (87.60 to 88.60): VN



#35

Dibromofluoromethane

Concen: 50.000 ug/l

RT: 7.59 min Scan# 1810

Delta R.T. -0.00 min

Lab File: VN052871.D

Acq: 13 Dec 2018 18:14


Tgt Ion:113 Resp: 551531

Ion Ratio Lower Upper

113 100

111 102.8 83.0 124.4



192 22.5 17.5 26.3




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#43

Isopropyl Acetate

Concen: 3.041 ug/l

RT: 8.17 min Scan# 1989

Delta R.T. 0.01 min

Lab File: VN052871.D

Acq: 13 Dec 2018 18:14

Instrument :

MSVOA_N

ClientSampleId :

PB115614ZHE#02


Tot Ion: 43 Resp: 66343

Ion Ratio Lower Upper

43 100

61 15.6 22.9 34.3#




87 9.3 10.6 15.8#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 61.00 (60.70 to 61.70): VN0

Ion 87.00 (86.70 to 87.70): VN0



#50

Toluene-d8

Concen: N.D. ug/l

RT: 10.09 min Scan# 2587

Delta R.T. -0.00 min

Lab File: VN052871.D


Acq: 13 Dec 2018 18:14

Tgt Ion: 98 Resp: 2350133

Ion Ratio Lower Upper



98 100


100 63.8 51.6 77.4



Abundance Ion 98.00 (97.70 to 98.70): VN0

Ion 100.00 (99.70 to 100.70): VN0









#62
4-Bromofluorobenzene
Concen: N.D. ug/l
RT: 12.40 min Scan# 3306
Delta R.T. -0.00 min
Lab File: VN052871.D
Acq: 13 Dec 2018 18:14

Instrument :
MSVOA_N
ClientSampleId :
PB115614ZHE#02

Tot Ion: 95 Resp: 769435
Ion Ratio Lower Upper
95 100
174 88.3 0.0 178.8
176 85.4 0.0 173.8




Abundance Ion 94.90 (94.60 to 95.60): VN052827.D (-3290) (-)
Ion 173.80 (173.50 to 174.50): VN052827.D (-3290) (-)
Ion 175.80 (175.50 to 176.50): VN052827.D (-3290) (-)






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. -0.00 min
Lab File: VN052871.D
Acq: 13 Dec 2018 18:14

Tgt Ion: 117 Resp: 1711322
Ion Ratio Lower Upper
117 100
82 54.0 42.8 64.2
119 32.3 25.5 38.3



Abundance Ion 116.90 (116.60 to 117.60): VN052827.D (-2981) (-)
Ion 82.00 (81.70 to 82.70): VN052827.D (-2981) (-)
Ion 118.90 (118.60 to 119.60): VN052827.D (-2981) (-)





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.35 min Scan# 3599

Delta R.T. -0.00 min

Lab File: VN052871.D

Acq: 13 Dec 2018 18:14

Instrument :

MSVOA_N

ClientSampleId :

PB115614ZHE#02


Tot Ion:152 Resp: 683753

Ion Ratio Lower Upper

152 100

115 56.7 28.3 85.0

150 158.8 0.0 347.8



Abundance

Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

