

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121321\
 Data File : VN070030.D
 Acq On : 13 Dec 2021 12:26
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Dec 14 05:26:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N121321W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Dec 14 00:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.662	128	155182	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.968	114	847703	30.000	ug/l	# 0.00
57) Chlorobenzene-d5	11.746	117	733819	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.440	65	424359	30.857	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	102.867%
60) 4-Bromofluorobenzene	12.731	95	302331	25.038	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	83.467%
63) Toluene-d8	10.443	98	1036296	30.559	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	101.867%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121321\
Data File : VN070030.D
Acq On : 13 Dec 2021 12:26
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Dec 14 05:26:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N121321W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Dec 14 00:59:04 2021
Response via : Initial Calibration

