

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121321\
 Data File : VN070036.D
 Acq On : 13 Dec 2021 16:54
 Operator : JC/MD
 Sample : M4986-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 001-WILLETS-PT-BLVD(DEC)

Quant Time: Dec 14 05:27:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N121321W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Dec 14 00:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.659	128	160850	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.968	114	865106	30.000	ug/l	# 0.00
57) Chlorobenzene-d5	11.747	117	770003	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.440	65	434165	30.458	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	101.533%
60) 4-Bromofluorobenzene	12.733	95	321439	25.370	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	84.567%
63) Toluene-d8	10.443	98	1065486	29.943	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.800%
Target Compounds						
15) Acetone	4.296	58	214164	90.433	ug/l	Qvalue 77
91) Naphthalene	15.507	128	39894	9.313	ug/l	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121321\
Data File : VN070036.D
Acq On : 13 Dec 2021 16:54
Operator : JC/MD
Sample : M4986-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
001-WILLETS-PT-BLVD(DEC)

Quant Time: Dec 14 05:27:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N121321W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Dec 14 00:59:04 2021
Response via : Initial Calibration

