

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121321\
 Data File : VN070037.D
 Acq On : 13 Dec 2021 17:20
 Operator : JC/MD
 Sample : M4986-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 002-35TH-AVE(DEC)

Quant Time: Dec 14 05:27:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N121321W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Dec 14 00:59:04 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.662	128	161497	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.968	114	860864	30.000	ug/l	# 0.00
57) Chlorobenzene-d5	11.747	117	762455	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.440	65	436387	30.491	ug/l	0.00
Spiked Amount	30.000	Range 91 - 110	Recovery	=	101.633%	
60) 4-Bromofluorobenzene	12.734	95	315178	25.122	ug/l	0.00
Spiked Amount	30.000	Range 63 - 112	Recovery	=	83.733%	
63) Toluene-d8	10.443	98	1064370	30.208	ug/l	0.00
Spiked Amount	30.000	Range 91 - 112	Recovery	=	100.700%	
Target Compounds						
					Qvalue	
15) Acetone	4.291	58	192400	80.918	ug/l	81
30) 2-Butanone	7.345	43	18079	1.773	ug/l	# 75
91) Naphthalene	15.507	128	14811	8.507	ug/l	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121321\
Data File : VN070037.D
Acq On : 13 Dec 2021 17:20
Operator : JC/MD
Sample : M4986-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
002-35TH-AVE(DEC)

Quant Time: Dec 14 05:27:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N121321W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Dec 14 00:59:04 2021
Response via : Initial Calibration

