

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121321\
 Data File : VN070040.D
 Acq On : 13 Dec 2021 18:36
 Operator : JC/MD
 Sample : M5061-01 5X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EFF-WW

Quant Time: Dec 14 05:28:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N121321W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Dec 14 00:59:04 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.659	128	159902	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.968	114	841932	30.000	ug/l	# 0.00
57) Chlorobenzene-d5	11.744	117	749009	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.440	65	434761	30.681	ug/l	0.00
Spiked Amount	30.000	Range 91 - 110	Recovery	= 102.267%		
60) 4-Bromofluorobenzene	12.734	95	309085	25.078	ug/l	0.00
Spiked Amount	30.000	Range 63 - 112	Recovery	= 83.600%		
63) Toluene-d8	10.443	98	1038467	30.002	ug/l	0.00
Spiked Amount	30.000	Range 91 - 112	Recovery	= 100.000%		
Target Compounds						
						Qvalue
15) Acetone	4.291	58	1318697	560.138	ug/l	76
25) Chloroform	7.823	83	34399	1.672	ug/l	93
58) 4-Methyl-2-Pentanone	10.336	43	18486	1.075	ug/l	# 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121321\
Data File : VN070040.D
Acq On : 13 Dec 2021 18:36
Operator : JC/MD
Sample : M5061-01 5X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EFF-WW

Quant Time: Dec 14 05:28:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N121321W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Dec 14 00:59:04 2021
Response via : Initial Calibration

