

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121321\
 Data File : VN070044.D
 Acq On : 13 Dec 2021 21:09
 Operator : JC/MD
 Sample : VN1213WBL02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1213WBL02

Quant Time: Dec 14 07:28:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	909213	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1704993	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1975312	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	784343	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	732262	52.162	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 104.320%	
35) Dibromofluoromethane	8.024	113	535586	50.150	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.300%	
50) Toluene-d8	10.443	98	2253408	52.548	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.100%	
62) 4-Bromofluorobenzene	12.731	95	951069	61.219	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 122.440%	
Target Compounds						Qvalue
93) 1,2,4-Trichlorobenzene	15.268	180	20445	4.375	ug/l	95
95) Naphthalene	15.509	128	95605	6.281	ug/l	99
96) 1,2,3-Trichlorobenzene	15.700	180	25998	4.408	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121321\
Data File : VN070044.D
Acq On : 13 Dec 2021 21:09
Operator : JC/MD
Sample : VN1213WBL02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1213WBL02

Quant Time: Dec 14 07:28:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

