

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121321\
 Data File : VN070053.D
 Acq On : 14 Dec 2021 1:24
 Operator : JC/MD
 Sample : M5007-09
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TOLL-MW-08

Quant Time: Dec 14 07:30:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1060469	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2024170	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2399839	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	965288	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	863546	52.740	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 105.480%	
35) Dibromofluoromethane	8.024	113	640265	50.498	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 101.000%	
50) Toluene-d8	10.443	98	2714003	53.310	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 106.620%	
62) 4-Bromofluorobenzene	12.731	95	1152423	62.483	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 124.960%	
Target Compounds						Qvalue
16) Acetone	4.293	43	24766	2.514	ug/l	91
19) Methyl tert-butyl Ether	5.621	73	19560	0.498	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121321\
Data File : VN070053.D
Acq On : 14 Dec 2021 1:24
Operator : JC/MD
Sample : M5007-09
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TOLL-MW-08

Quant Time: Dec 14 07:30:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

