

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121321\
 Data File : VN070055.D
 Acq On : 14 Dec 2021 2:15
 Operator : JC/MD
 Sample : M5007-11
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TOLL-MW-10

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/14/2021
 Supervised By : Amit Patel 12/15/2021

Quant Time: Dec 14 07:30:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1076739	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2094322	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2456544	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	995566	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	883381	53.136	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 106.280%			
35) Dibromofluoromethane	8.024	113	640732	48.842	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 97.680%			
50) Toluene-d8	10.440	98	2809893	53.344	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 106.680%			
62) 4-Bromofluorobenzene	12.731	95	1179004	61.783	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 123.560%			
Target Compounds						Qvalue
4) Vinyl Chloride	2.451	62	101376	7.007	ug/l	91
16) Acetone	4.299	43	32173m	3.217	ug/l	
19) Methyl tert-butyl Ether	5.624	73	129635	3.251	ug/l #	79
27) cis-1,2-Dichloroethene	7.327	96	88642	6.586	ug/l	96
44) Trichloroethene	9.217	130	3051	0.210	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121321\
Data File : VN070055.D
Acq On : 14 Dec 2021 2:15
Operator : JC/MD
Sample : M5007-11
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TOLL-MW-10

Quant Time: Dec 14 07:30:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 12/14/2021
Supervised By : Amit Patel 12/15/2021

