

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN121321\  
 Data File : VN070056.D  
 Acq On : 14 Dec 2021 2:41  
 Operator : JC/MD  
 Sample : M5007-13  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 36 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 TB-2021.12.08

Quant Time: Dec 14 07:31:09 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N120221W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Dec 03 08:46:47 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units      | Dev(Min)  |
|-----------------------------|--------|-------|----------|----------|------------|-----------|
| -----                       |        |       |          |          |            |           |
| Internal Standards          |        |       |          |          |            |           |
| 1) Pentafluorobenzene       | 8.086  | 168   | 1065860  | 50.000   | ug/l       | 0.00      |
| 34) 1,4-Difluorobenzene     | 8.968  | 114   | 2060228  | 50.000   | ug/l       | 0.00      |
| 63) Chlorobenzene-d5        | 11.744 | 117   | 2432335  | 50.000   | ug/l       | 0.00      |
| 72) 1,4-Dichlorobenzene-d4  | 13.675 | 152   | 997312   | 50.000   | ug/l       | 0.00      |
| System Monitoring Compounds |        |       |          |          |            |           |
| 33) 1,2-Dichloroethane-d4   | 8.440  | 65    | 880064   | 53.477   | ug/l       | 0.00      |
| Spiked Amount               | 50.000 | Range | 61 - 141 | Recovery | = 106.960% |           |
| 35) Dibromofluoromethane    | 8.024  | 113   | 642389   | 49.779   | ug/l       | 0.00      |
| Spiked Amount               | 50.000 | Range | 69 - 133 | Recovery | = 99.560%  |           |
| 50) Toluene-d8              | 10.440 | 98    | 2753702  | 53.143   | ug/l       | 0.00      |
| Spiked Amount               | 50.000 | Range | 65 - 126 | Recovery | = 106.280% |           |
| 62) 4-Bromofluorobenzene    | 12.733 | 95    | 1193298  | 63.567   | ug/l       | 0.00      |
| Spiked Amount               | 50.000 | Range | 58 - 135 | Recovery | = 127.140% |           |
| Target Compounds            |        |       |          |          |            |           |
| 16) Acetone                 | 4.293  | 43    | 22928    | 2.316    | ug/l       | Qvalue 90 |
| -----                       |        |       |          |          |            |           |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN121321\  
Data File : VN070056.D  
Acq On : 14 Dec 2021 2:41  
Operator : JC/MD  
Sample : M5007-13  
Misc : 5.00mL/MSVOA\_N/WATER  
ALS Vial : 36 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
TB-2021.12.08

Quant Time: Dec 14 07:31:09 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N120221W.M  
Quant Title : SW846 8260  
QLast Update : Fri Dec 03 08:46:47 2021  
Response via : Initial Calibration

