

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121321\
 Data File : VN070043.D
 Acq On : 13 Dec 2021 20:18
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 14 07:27:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	85	0.00
2 T	Dichlorodifluoromethane	0.583	0.543	6.9	70	0.00
3 P	Chloromethane	0.731	0.665	9.0	74	0.00
4 C	Vinyl Chloride	0.672	0.623	7.3#	73	0.00
5 T	Bromomethane	0.365	0.345	5.5	74	0.00
6 T	Chloroethane	0.412	0.372	9.7	72	0.00
7 T	Trichlorofluoromethane	0.882	0.823	6.7	74	0.00
8 T	Diethyl Ether	0.340	0.342	-0.6	79	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.494	0.496	-0.4	80	0.00
10 T	Methyl Iodide	0.675	0.669	0.9	75	0.00
11 T	Tert butyl alcohol	0.146	0.127	13.0	68	0.00
12 CM	1,1-Dichloroethene	0.476	0.464	2.5#	78	0.00
13 T	Acrolein	0.124	0.116	6.5	79	0.00
14 T	Allyl chloride	1.073	1.103	-2.8	82	0.00
15 T	Acrylonitrile	0.393	0.390	0.8	77	0.00
16 T	Acetone	0.464	0.416	10.3	70	0.00
17 T	Carbon Disulfide	1.230	1.261	-2.5	80	0.00
18 T	Methyl Acetate	1.402	1.330	5.1	77	0.00
19 T	Methyl tert-butyl Ether	1.852	1.834	1.0	78	0.00
20 T	Methylene Chloride	0.617	0.586	5.0	80	0.00
21 T	trans-1,2-Dichloroethene	0.517	0.498	3.7	77	0.00
22 T	Diisopropyl ether	2.098	2.195	-4.6	81	0.00
23 T	Vinyl Acetate	1.707	1.576	7.7	69	0.00
24 P	1,1-Dichloroethane	1.087	1.112	-2.3	81	0.00
25 T	2-Butanone	0.631	0.593	6.0	73	0.00
26 T	2,2-Dichloropropane	0.845	0.804	4.9	76	0.00
27 T	cis-1,2-Dichloroethene	0.625	0.619	1.0	78	0.00
28 T	Bromochloromethane	0.565	0.601	-6.4	85	0.00
29 T	Tetrahydrofuran	0.383	0.375	2.1	75	0.00
30 C	Chloroform	1.097	1.119	-2.0#	81	0.00
31 T	Cyclohexane	1.074	1.000	6.9	76	0.00
32 T	1,1,1-Trichloroethane	0.935	0.943	-0.9	78	0.00
33 S	1,2-Dichloroethane-d4	0.772	0.776	-0.5	81	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	84	0.00
35 S	Dibromofluoromethane	0.313	0.331	-5.8	82	0.00
36 T	1,1-Dichloropropene	0.476	0.480	-0.8	77	0.00
37 T	Ethyl Acetate	0.708	0.673	4.9	73	0.00
38 T	Carbon Tetrachloride	0.467	0.496	-6.2	81	0.00
39 T	Methylcyclohexane	0.566	0.552	2.5	73	0.00
40 TM	Benzene	1.455	1.465	-0.7	78	0.00
41 T	Methacrylonitrile	0.333	0.344	-3.3	77	0.00
42 TM	1,2-Dichloroethane	0.587	0.588	-0.2	78	0.00
43 T	Isopropyl Acetate	1.048	1.053	-0.5	76	0.00
44 TM	Trichloroethene	0.347	0.342	1.4	76	0.00
45 C	1,2-Dichloropropane	0.391	0.404	-3.3#	80	0.00
46 T	Dibromomethane	0.256	0.264	-3.1	79	0.00
47 T	Bromodichloromethane	0.485	0.557	-14.8	86	0.00
48 T	Methyl methacrylate	0.489	0.494	-1.0	76	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121321\
 Data File : VN070043.D
 Acq On : 13 Dec 2021 20:18
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 14 07:27:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.008	11.1	71	0.00
50 S	Toluene-d8	1.258	1.250	0.6	77	0.00
51 T	4-Methyl-2-Pentanone	0.680	0.714	-5.0	77	0.00
52 CM	Toluene	0.888	0.894	-0.7#	76	0.00
53 T	t-1,3-Dichloropropene	0.510	0.564	-10.6	80	0.00
54 T	cis-1,3-Dichloropropene	0.548	0.608	-10.9	79	0.00
55 T	1,1,2-Trichloroethane	0.358	0.372	-3.9	79	0.00
56 T	Ethyl methacrylate	0.608	0.618	-1.6	75	0.00
57 T	1,3-Dichloropropane	0.631	0.648	-2.7	78	0.00
58 T	2-Chloroethyl Vinyl ether	0.154	0.177	-14.9	84	0.00
59 T	2-Hexanone	0.514	0.520	-1.2	74	0.00
60 T	Dibromochloromethane	0.340	0.406	-19.4	86	0.00
61 T	1,2-Dibromoethane	0.371	0.381	-2.7	77	0.00
62 S	4-Bromofluorobenzene	0.456	0.467	-2.4	80	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	84	0.00
64 T	Tetrachloroethene	0.339	0.312	8.0	71	0.00
65 PM	Chlorobenzene	1.014	1.006	0.8	76	0.00
66 T	1,1,1,2-Tetrachloroethane	0.357	0.385	-7.8	80	0.00
67 C	Ethyl Benzene	1.843	1.899	-3.0#	77	0.00
68 T	m/p-Xylenes	0.671	0.693	-3.3	76	0.00
69 T	o-Xylene	0.676	0.690	-2.1	76	0.00
70 T	Styrene	1.072	1.152	-7.5	78	0.00
71 P	Bromoform	0.264	0.334	-26.5#	90	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
73 T	Isopropylbenzene	4.449	4.296	3.4	77	0.00
74 T	N-amyl acetate	1.967	1.909	2.9	74	0.00
75 P	1,1,2,2-Tetrachloroethane	1.538	1.501	2.4	79	0.00
76 T	1,2,3-Trichloropropane	1.434	1.344	6.3	74	0.00
77 T	Bromobenzene	1.045	0.998	4.5	77	0.00
78 T	n-propylbenzene	4.950	4.930	0.4	77	0.00
79 T	2-Chlorotoluene	3.128	3.027	3.2	77	0.00
80 T	1,3,5-Trimethylbenzene	3.581	3.531	1.4	77	0.00
81 T	trans-1,4-Dichloro-2-butene	0.396	0.435	-9.8	80	0.00
82 T	4-Chlorotoluene	2.942	2.918	0.8	78	0.00
83 T	tert-Butylbenzene	3.171	3.077	3.0	76	0.00
84 T	1,2,4-Trimethylbenzene	3.410	3.426	-0.5	76	0.00
85 T	sec-Butylbenzene	4.301	4.305	-0.1	76	0.00
86 T	p-Isopropyltoluene	3.362	3.412	-1.5	76	0.00
87 T	1,3-Dichlorobenzene	1.717	1.698	1.1	77	0.00
88 T	1,4-Dichlorobenzene	1.687	1.634	3.1	77	0.00
89 T	n-Butylbenzene	2.727	2.781	-2.0	76	0.00
90 T	Hexachloroethane	0.590	0.678	-14.9	86	0.00
91 T	1,2-Dichlorobenzene	1.673	1.646	1.6	77	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.278	0.291	-4.7	80	0.00
93 T	1,2,4-Trichlorobenzene	0.760	0.752	1.1	76	0.00
94 T	Hexachlorobutadiene	0.443	0.445	-0.5	77	0.00
95 T	Naphthalene	2.199	2.108	4.1	74	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121321\
Data File : VN070043.D
Acq On : 13 Dec 2021 20:18
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Dec 14 07:27:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.749	0.758	-1.2	77	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6