

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121323\
 Data File : VN080456.D
 Acq On : 13 Dec 2023 12:55
 Operator : JC\MD
 Sample : 05688-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Dec 14 04:05:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	205772	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	377626	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	390457	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	173259	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	164065	62.425	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 124.860%	
35) Dibromofluoromethane	8.165	113	138551	57.867	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 115.740%	
50) Toluene-d8	10.565	98	497959	55.766	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 111.540%	
62) 4-Bromofluorobenzene	12.847	95	199306	58.779	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 117.560%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121323\
Data File : VN080456.D
Acq On : 13 Dec 2023 12:55
Operator : JC\MD
Sample : 05688-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Dec 14 04:05:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

