

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121323\
 Data File : VN080458.D
 Acq On : 13 Dec 2023 13:43
 Operator : JC\MD
 Sample : 05688-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PMW-1

Quant Time: Dec 14 04:06:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	187700	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	344607	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	358667	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	157964	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	150372	62.724	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 125.440%#	
35) Dibromofluoromethane	8.165	113	128539	58.829	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 117.660%	
50) Toluene-d8	10.565	98	452718	55.557	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 111.120%	
62) 4-Bromofluorobenzene	12.847	95	183016	59.147	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 118.300%	
Target Compounds						
16) Acetone	4.442	43	2547	5.992	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121323\
Data File : VN080458.D
Acq On : 13 Dec 2023 13:43
Operator : JC\MD
Sample : 05688-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PMW-1

Quant Time: Dec 14 04:06:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

