

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121323\
 Data File : VN080459.D
 Acq On : 13 Dec 2023 14:06
 Operator : JC\MD
 Sample : 05688-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-7

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/14/2023
 Supervised By : Mahesh Dadoda 12/14/2023

Quant Time: Dec 14 04:06:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	196806	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	364434	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	388426	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	175264	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	162526	64.657	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 129.320%#			
35) Dibromofluoromethane	8.165	113	137138	59.350	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 118.700%			
50) Toluene-d8	10.565	98	490221	56.886	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 113.780%#			
62) 4-Bromofluorobenzene	12.853	95	195758	59.823	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 119.640%			
Target Compounds					Qvalue	
30) Chloroform	7.965	83	5127	1.172	ug/l	90
64) Tetrachloroethene	11.094	164	1387	0.462	ug/l #	59

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121323\
Data File : VN080459.D
Acq On : 13 Dec 2023 14:06
Operator : JC\MD
Sample : 05688-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-7

Quant Time: Dec 14 04:06:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 12/14/2023
Supervised By :Mahesh Dadoda 12/14/2023

