

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121323\
 Data File : VN080468.D
 Acq On : 13 Dec 2023 17:39
 Operator : JC\MD
 Sample : 05791-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1B-K-T

Quant Time: Dec 14 04:10:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	254274	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	458129	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	469709	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	214876	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	165201	50.868	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.740%
35) Dibromofluoromethane	8.165	113	139432	48.002	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.000%
50) Toluene-d8	10.565	98	492714	45.482	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.960%
62) 4-Bromofluorobenzene	12.847	95	201033	48.870	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.740%
Target Compounds						
						Qvalue
16) Acetone	4.424	43	10211	17.732	ug/l #	79
20) Methylene Chloride	5.283	84	2969	0.963	ug/l #	93
30) Chloroform	7.965	83	101866	18.020	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121323\
Data File : VN080468.D
Acq On : 13 Dec 2023 17:39
Operator : JC\MD
Sample : 05791-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1B-K-T

Quant Time: Dec 14 04:10:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

