

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121323\
 Data File : VN080469.D
 Acq On : 13 Dec 2023 18:03
 Operator : JC\MD
 Sample : 05791-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1B-U-DD

Quant Time: Dec 14 04:10:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	249067	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	448453	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	460407	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	208552	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	159743	50.215	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.440%	
35) Dibromofluoromethane	8.165	113	135703	47.726	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.460%	
50) Toluene-d8	10.571	98	477718	45.049	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 90.100%	
62) 4-Bromofluorobenzene	12.847	95	196325	48.755	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.520%	
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	11755	20.840	ug/l	97
19) Methyl tert-butyl Ether	5.795	73	35685	3.995	ug/l	99
20) Methylene Chloride	5.265	84	3597	1.192	ug/l #	72
30) Chloroform	7.971	83	116079	20.963	ug/l	91
47) Bromodichloromethane	9.894	83	3759	0.775	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121323\
Data File : VN080469.D
Acq On : 13 Dec 2023 18:03
Operator : JC\MD
Sample : 05791-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1B-U-DD

Quant Time: Dec 14 04:10:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

