

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121323\
 Data File : VN080470.D
 Acq On : 13 Dec 2023 18:26
 Operator : JC\MD
 Sample : 05791-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1B-00-XX

Quant Time: Dec 14 04:11:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	246797	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	440446	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	448727	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	204541	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	161500	51.235	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.460%
35) Dibromofluoromethane	8.171	113	136857	49.007	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.020%
50) Toluene-d8	10.571	98	476557	45.757	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.520%
62) 4-Bromofluorobenzene	12.853	95	192081	48.569	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.140%
Target Compounds						
						Qvalue
16) Acetone	4.424	43	12892	23.066	ug/l #	80
20) Methylene Chloride	5.265	84	3036	1.015	ug/l #	85
25) 2-Butanone	7.488	43	2409	2.237	ug/l	93
30) Chloroform	7.965	83	100556	18.327	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121323\
Data File : VN080470.D
Acq On : 13 Dec 2023 18:26
Operator : JC\MD
Sample : 05791-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1B-00-XX

Quant Time: Dec 14 04:11:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

