

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121323\
 Data File : VN080471.D
 Acq On : 13 Dec 2023 18:50
 Operator : JC\MD
 Sample : 05791-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1B-A-J

Quant Time: Dec 14 04:11:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	258185	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	461189	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	474156	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	211848	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	167416	50.769	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.540%
35) Dibromofluoromethane	8.165	113	138446	47.346	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.700%
50) Toluene-d8	10.570	98	489010	44.841	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.680%
62) 4-Bromofluorobenzene	12.853	95	197637	47.726	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.460%
Target Compounds						
						Qvalue
16) Acetone	4.430	43	10613	18.151	ug/l	98
20) Methylene Chloride	5.271	84	2984	0.954	ug/l #	82
30) Chloroform	7.965	83	125392	21.845	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121323\
Data File : VN080471.D
Acq On : 13 Dec 2023 18:50
Operator : JC\MD
Sample : 05791-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1B-A-J

Quant Time: Dec 14 04:11:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

