

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121323\
 Data File : VN080472.D
 Acq On : 13 Dec 2023 19:14
 Operator : JC\MD
 Sample : 05791-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1B-EE-NN

Quant Time: Dec 14 04:12:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	241424	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	436504	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	435649	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	190354	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	158575	51.426	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.860%
35) Dibromofluoromethane	8.171	113	131281	47.435	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.860%
50) Toluene-d8	10.570	98	461709	44.732	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.460%
62) 4-Bromofluorobenzene	12.847	95	182959	46.680	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.360%
Target Compounds						
						Qvalue
16) Acetone	4.430	43	10823	19.795	ug/l #	85
19) Methyl tert-butyl Ether	5.800	73	12361	1.428	ug/l	92
20) Methylene Chloride	5.277	84	4732	1.617	ug/l #	90
30) Chloroform	7.971	83	166644	31.048	ug/l	92
47) Bromodichloromethane	9.888	83	9883	2.092	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121323\
Data File : VN080472.D
Acq On : 13 Dec 2023 19:14
Operator : JC\MD
Sample : 05791-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1B-EE-NN

Quant Time: Dec 14 04:12:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

