

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN121323\  
 Data File : VN080473.D  
 Acq On : 13 Dec 2023 19:38  
 Operator : JC\MD  
 Sample : 05812-01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 25 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 RW2

Quant Time: Dec 14 04:12:31 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N121223W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Dec 13 01:57:14 2023  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|--------|----------|
| -----                       |        |       |          |          |        |          |
| Internal Standards          |        |       |          |          |        |          |
| 1) Pentafluorobenzene       | 8.224  | 168   | 229445   | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 9.106  | 114   | 412998   | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 11.865 | 117   | 431793   | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.794 | 152   | 193263   | 50.000   | ug/l   | 0.00     |
| System Monitoring Compounds |        |       |          |          |        |          |
| 33) 1,2-Dichloroethane-d4   | 8.583  | 65    | 151298   | 51.628   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | =      | 103.260% |
| 35) Dibromofluoromethane    | 8.171  | 113   | 126511   | 48.313   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =      | 96.620%  |
| 50) Toluene-d8              | 10.565 | 98    | 454724   | 46.562   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | =      | 93.120%  |
| 62) 4-Bromofluorobenzene    | 12.847 | 95    | 185406   | 49.997   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =      | 100.000% |
| Target Compounds            |        |       |          |          |        |          |
|                             |        |       |          |          | Qvalue |          |
| 16) Acetone                 | 4.436  | 43    | 3976     | 7.652    | ug/l # | 85       |
| 65) Chlorobenzene           | 11.894 | 112   | 186101   | 19.375   | ug/l   | 99       |
| -----                       |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN121323\  
Data File : VN080473.D  
Acq On : 13 Dec 2023 19:38  
Operator : JC\MD  
Sample : 05812-01  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 25 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
RW2

Quant Time: Dec 14 04:12:31 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N121223W.M  
Quant Title : SW846 8260  
QLast Update : Wed Dec 13 01:57:14 2023  
Response via : Initial Calibration

