

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121324\
 Data File : VN085191.D
 Acq On : 13 Dec 2024 12:53
 Operator : JC\MD
 Sample : VN1213WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1213WBL01

Quant Time: Dec 13 23:30:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 23 02:54:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	151380	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	264382	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	226403	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	89452	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	109015	50.557	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.120%
35) Dibromofluoromethane	8.165	113	89196	50.029	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.060%
50) Toluene-d8	10.565	98	310576	48.607	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.220%
62) 4-Bromofluorobenzene	12.847	95	100184	41.927	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	83.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121324\
Data File : VN085191.D
Acq On : 13 Dec 2024 12:53
Operator : JC\MD
Sample : VN1213WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1213WBL01

Quant Time: Dec 13 23:30:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 23 02:54:10 2024
Response via : Initial Calibration

