

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121324\
 Data File : VN085194.D
 Acq On : 13 Dec 2024 14:16
 Operator : JC\MD
 Sample : P5270-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW8-20241211

Quant Time: Dec 13 23:33:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 23 02:54:10 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	134343	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	233695	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	202190	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	81706	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	97217	50.803	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.600%
35) Dibromofluoromethane	8.165	113	81390	51.645	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.280%
50) Toluene-d8	10.565	98	275639	48.804	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.600%
62) 4-Bromofluorobenzene	12.847	95	93207	44.130	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	88.260%
Target Compounds						Qvalue
30) Chloroform	7.965	83	52513	17.051	ug/l	97
64) Tetrachloroethene	11.100	164	923	0.677	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121324\
Data File : VN085194.D
Acq On : 13 Dec 2024 14:16
Operator : JC\MD
Sample : P5270-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW8-20241211

Quant Time: Dec 13 23:33:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 23 02:54:10 2024
Response via : Initial Calibration

