

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121324\
 Data File : VN085197.D
 Acq On : 13 Dec 2024 15:28
 Operator : JC\MD
 Sample : P5270-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW17-20241211

Quant Time: Dec 13 23:35:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 23 02:54:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.223	168	137054	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	242188	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.864	117	210970	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	83538	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	104535	53.546	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.100%
35) Dibromofluoromethane	8.165	113	85236	52.188	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.380%
50) Toluene-d8	10.565	98	282126	48.201	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.400%
62) 4-Bromofluorobenzene	12.847	95	95621	43.685	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	87.360%
Target Compounds						Qvalue
30) Chloroform	7.959	83	37777	12.024	ug/l	97
64) Tetrachloroethene	11.106	164	826	0.581	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121324\
Data File : VN085197.D
Acq On : 13 Dec 2024 15:28
Operator : JC\MD
Sample : P5270-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW17-20241211

Quant Time: Dec 13 23:35:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 23 02:54:10 2024
Response via : Initial Calibration

