

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121418\
 Data File : VN052908.D
 Acq On : 14 Dec 2018 23:59
 Operator : MD\SY
 Sample : J6403-56
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 33 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Dec 15 03:04:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 13 01:08:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1346966	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2085470	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1816869	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	682850	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	684527	47.79	ug/l	0.00
Spiked Amount			Recovery	=	95.58%	
35) Dibromofluoromethane	7.59	113	555505	47.72	ug/l	0.00
Spiked Amount			Recovery	=	95.44%	
50) Toluene-d8	10.09	98	2525810	50.04	ug/l	0.00
Spiked Amount			Recovery	=	100.08%	
62) 4-Bromofluorobenzene	12.40	95	787573	46.05	ug/l	0.00
Spiked Amount			Recovery	=	92.10%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
8) Diethyl Ether	3.42	74	2967	0.394	ug/l	81
14) Allyl chloride	4.56	41	7530	0.372	ug/l #	24
16) Acetone	3.82	43	33393	10.381	ug/l	98
18) Methyl Acetate	4.33	43	5825	0.493	ug/l #	78
20) Methylene Chloride	4.55	84	191896	13.256	ug/l	97
29) Tetrahydrofuran	7.22	42	1931	0.567	ug/l #	63
31) Cyclohexane	7.67	56	21707	0.318	ug/l #	1
36) 1,1-Dichloropropene	7.67	75	90509	4.713	ug/l #	48
38) Carbon Tetrachloride	7.67	117	116426	6.586	ug/l #	16
43) Isopropyl Acetate	8.17	43	69147	2.923	ug/l #	78
45) 1,2-Dichloropropane	9.27	63	3504	0.232	ug/l #	44
48) Methyl methacrylate	9.18	41	15018	1.448	ug/l #	12
49) 1,4-Dioxane	9.27	88	193	1.639	ug/l #	1
51) 4-Methyl-2-Pentanone	10.03	43	17923	1.652	ug/l #	35
54) cis-1,3-Dichloropropene	9.90	75	126893	5.828	ug/l #	61
57) 1,3-Dichloropropane	10.76	76	24911	1.166	ug/l #	42
59) 2-Hexanone	10.77	43	331298	44.571	ug/l #	50
68) m/p-Xylenes	11.62	106	8556	0.352	ug/l	98
71) Bromoform	12.40	173	4876	0.547	ug/l #	1
74) N-amyl acetate	11.88	43	37227	2.660	ug/l #	52
76) 1,2,3-Trichloropropane	12.40	75	369447	36.671	ug/l #	100
80) 1,3,5-Trimethylbenzene	12.73	105	15434	0.375	ug/l	97
81) trans-1,4-Dichloro-2-buten	12.40	75	369447	117.805	ug/l #	12
84) 1,2,4-Trimethylbenzene	13.04	105	27640	0.663	ug/l	99
95) Naphthalene	15.13	128	38520	1.543	ug/l	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121418\
Data File : VN052908.D
Acq On : 14 Dec 2018 23:59
Operator : MD\SY
Sample : J6403-56
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 33 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Dec 15 03:04:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 13 01:08:35 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. -0.00 min

Lab File: VN052908.D

Acq: 14 Dec 2018 23:59


Instrument :
MSVOA_N
ClientSampled :

Tot Ion:168 Resp: 1346966

Ion Ratio Lower Upper

168 100

99 50.6 40.2 60.2



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN

7.67

600000

500000

400000

300000



200000

100000

0

Time-->

7.60 7.70 7.80



#8

Diethyl Ether

Concen: 0.394 ug/l

RT: 3.42 min Scan# 512

Delta R.T. 0.01 min

Lab File: VN052908.D


Acq: 14 Dec 2018 23:59

Tgt Ion: 74 Resp: 2967

Ion Ratio Lower Upper

74 100

45 117.9 49.4 148.0



Abundance Ion 74.00 (73.70 to 74.70): VN

Ion 45.00 (44.70 to 45.70): VN

2000

1500


1000


500

0

Time-->

3.35 3.40 3.45









#14
 Allvl chloride
 Concen: 0.372 ug/l
 RT: 4.56 min Scan# 868
 Delta R.T. 0.23 min
 Lab File: VN052908.D
 Acq: 14 Dec 2018 23:59




Instrument :
 MSVOA_N
 ClientSampled :


Tot Ion: 41 Resp: 7530
 Ion Ratio Lower Upper
 41 100
 39 0.0 54.1 81.1#
 76 0.0 30.3 45.5#



#16
 Acetone
 Concen: 10.381 ug/l
 RT: 3.82 min Scan# 636
 Delta R.T. 0.00 min
 Lab File: VN052908.D
 Acq: 14 Dec 2018 23:59

Tgt Ion: 43 Resp: 33393
 Ion Ratio Lower Upper
 43 100
 58 30.4 25.4 38.0









#18
Methvl Acetate
Concen: 0.493 ug/l
RT: 4.33 min Scan# 795
Delta R.T. 0.01 min
Lab File: VN052908.D
Acq: 14 Dec 2018 23:59




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion: 43 Resp: 5825
Ion Ratio Lower Upper
43 100
74 13.2 19.2 28.8#



#20
Methylene Chloride
Concen: 13.256 ug/l
RT: 4.55 min Scan# 865
Delta R.T. -0.00 min
Lab File: VN052908.D
Acq: 14 Dec 2018 23:59

Tgt Ion: 84 Resp: 191896
Ion Ratio Lower Upper
84 100
49 116.7 96.8 145.2
51 35.8 29.9 44.9
86 64.3 51.8 77.6









#29
Tetrahydrofuran
Concen: 0.567 ug/l
RT: 7.22 min Scan# 1693
Delta R.T. 0.01 min
Lab File: VN052908.D
Acq: 14 Dec 2018 23:59




Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 42 Resp: 1931
Ion Ratio Lower Upper
42 100
72 25.7 34.7 52.1#
71 11.2 32.3 48.5#



#31
Cyclohexane
Concen: 0.318 ug/l
RT: 7.67 min Scan# 1833
Delta R.T. 0.01 min
Lab File: VN052908.D
Acq: 14 Dec 2018 23:59

Tgt Ion: 56 Resp: 21707
Ion Ratio Lower Upper
56 100
69 182.9 26.6 39.8#
84 165.5 68.4 102.6#









#33
 1,2-Dichloroethane-d4
 Concen: Below ug/l
 RT: 8.03 min Scan# 1946
 Delta R.T. 0.00 min
 Lab File: VN052908.D
 Acq: 14 Dec 2018 23:59




Instrument :
 MSVOA_N
 ClientSampleId :


Tot Ion: 65 Resp: 684527
 Ion Ratio Lower Upper
 65 100
 67 54.6 0.0 110.4



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.59 min Scan# 2119
 Delta R.T. -0.00 min
 Lab File: VN052908.D
 Acq: 14 Dec 2018 23:59

Tgt Ion: 114 Resp: 2085470
 Ion Ratio Lower Upper
 114 100
 63 19.6 0.0 39.8
 88 15.4 0.0 31.0









#35
 Dibromofluoromethane
 Concen: 50.000 ug/l
 RT: 7.59 min Scan# 1810
 Delta R.T. -0.00 min
 Lab File: VN052908.D
 Acq: 14 Dec 2018 23:59




Instrument :
 MSVOA_N
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
113	100		
111	102.9	83.0	124.4
192	22.7	17.5	26.3



#36
 1,1-Dichloropropene
 Concen: 4.713 ug/l
 RT: 7.67 min Scan# 1833
 Delta R.T. -0.13 min
 Lab File: VN052908.D
 Acq: 14 Dec 2018 23:59

Tgt Ion	Ratio	Lower	Upper
75	100		
110	8.0	18.2	54.6#
77	0.0	24.9	37.3#





#38

Carbon Tetrachloride

Concen: 6.586 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. -0.11 min

Lab File: VN052908.D

Acq: 14 Dec 2018 23:59

Instrument :

MSVOA_N

ClientSampled :


Tot Ion:117 Resp: 116426

Ion Ratio Lower Upper

117 100

119 5.4 77.1 115.7#




121 0.0 24.9 37.3#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#43

Isopropyl Acetate

Concen: 2.923 ug/l

RT: 8.17 min Scan# 1989

Delta R.T. 0.01 min

Lab File: VN052908.D

Acq: 14 Dec 2018 23:59


Tgt Ion: 43 Resp: 69147

Ion Ratio Lower Upper

43 100

61 14.4 22.9 34.3#



87 8.5 10.6 15.8#




Abundance Ion 43.00 (42.70 to 43.70): V

Ion 61.00 (60.70 to 61.70): V

Ion 87.00 (86.70 to 87.70): V









#45
 1,2-Dichloropropane
 Concen: 0.232 ug/l
 RT: 9.27 min Scan# 2331
 Delta R.T. 0.15 min
 Lab File: VN052908.D
 Acq: 14 Dec 2018 23:59




Instrument :
 MSVOA_N
 ClientSampleId :


Tot Ion: 63 Resp: 3504
 Ion Ratio Lower Upper
 63 100
 65 0.0 24.6 37.0#



#48
 Methyl methacrylate
 Concen: 1.448 ug/l
 RT: 9.18 min Scan# 2304
 Delta R.T. -0.02 min
 Lab File: VN052908.D
 Acq: 14 Dec 2018 23:59

Tgt Ion: 41 Resp: 15018
 Ion Ratio Lower Upper
 41 100
 69 0.0 72.2 108.2#
 39 0.0 41.9 62.9#









#49
 1,4-Dioxane
 Concen: 1.639 ug/l
 RT: 9.27 min Scan# 2331
 Delta R.T. 0.07 min
 Lab File: VN052908.D
 Acq: 14 Dec 2018 23:59

Instrument :
 MSVOA_N
 ClientSampleId :

Tot Ion: 88 Resp: 193
 Ion Ratio Lower Upper
 88 100
 43 1361752.3 26.6 40.0#
 58 5694.8 58.4 87.6#




Abundance Ion 88.00 (87.70 to 88.70): VNC
 Ion 43.00 (42.70 to 43.70): VNC
 Ion 58.00 (57.70 to 58.70): VNC






#50
 Toluene-d8
 Concen: 1.639 ug/l
 RT: 10.09 min Scan# 2587
 Delta R.T. -0.00 min
 Lab File: VN052908.D
 Acq: 14 Dec 2018 23:59

Tgt Ion: 98 Resp: 2525810
 Ion Ratio Lower Upper
 98 100
 100 64.5 51.6 77.4



Abundance Ion 98.00 (97.70 to 98.70): VNC
 Ion 100.00 (99.70 to 100.70): VNC





#51

4-Methyl-2-Pentanone

Concen: 1.652 ug/l

RT: 10.03 min Scan# 2567

Delta R.T. 0.04 min

Lab File: VN052908.D

Acq: 14 Dec 2018 23:59

Instrument :

MSVOA_N





ClientSampled :

Tot Ion: 43 Resp: 17923

Ion Ratio Lower Upper

43 100

58 0.0 32.2 48.4#



#54

cis-1,3-Dichloropropene

Concen: 5.828 ug/l

RT: 9.90 min Scan# 2529

Delta R.T. 0.06 min

Lab File: VN052908.D

Acq: 14 Dec 2018 23:59




Tgt Ion: 75 Resp: 126893


Ion Ratio Lower Upper

75 100

77 0.8 25.8 38.8#

39 27.2 34.8 52.2#









#57
 1,3-Dichloropropane
 Concen: 1.166 ug/l
 RT: 10.76 min Scan# 2796
 Delta R.T. 0.05 min
 Lab File: VN052908.D
 Acq: 14 Dec 2018 23:59




Instrument :
 MSVOA_N
 ClientSampleId :


Tot Ion: 76 Resp: 24911
 Ion Ratio Lower Upper
 76 100
 78 0.0 25.8 38.8#



#59
 2-Hexanone
 Concen: 44.571 ug/l
 RT: 10.77 min Scan# 2797
 Delta R.T. 0.02 min
 Lab File: VN052908.D
 Acq: 14 Dec 2018 23:59

Tgt Ion: 43 Resp: 331298
 Ion Ratio Lower Upper
 43 100
 58 19.6 28.1 84.3#









#62
4-Bromofluorobenzene
Concen: N.D. ug/l
RT: 12.40 min Scan# 3306
Delta R.T. -0.00 min
Lab File: VN052908.D
Acq: 14 Dec 2018 23:59




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion: 95 Resp: 787573
Ion Ratio Lower Upper
95 100
174 89.8 0.0 178.8
176 87.5 0.0 173.8



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2997
Delta R.T. 0.00 min
Lab File: VN052908.D
Acq: 14 Dec 2018 23:59

Tgt Ion: 117 Resp: 1816869
Ion Ratio Lower Upper
117 100
82 53.3 42.8 64.2
119 32.1 25.5 38.3





#68

m/p-Xvlenes

Concen: 0.352 ug/l

RT: 11.62 min Scan# 3063

Delta R.T. -0.00 min

Lab File: VN052908.D

Acq: 14 Dec 2018 23:59

Instrument :

MSVOA_N


ClientSampled :

Tot Ion:106 Resp: 8556

Ion Ratio Lower Upper




106 100

91 199.3 162.2 243.4



Abundance Ion 106.00 (105.70 to 106.70): V

Ion 91.00 (90.70 to 91.70): VN



#71

Bromoform

Concen: 0.547 ug/l

RT: 12.40 min Scan# 3306

Delta R.T. 0.28 min

Lab File: VN052908.D

Acq: 14 Dec 2018 23:59


Tgt Ion:173 Resp: 4876

Ion Ratio Lower Upper

173 100

175 878.8 24.3 72.9#



254 0.0 0.1 0.1#




Abundance Ion 172.70 (172.40 to 173.40): V

Ion 174.70 (174.40 to 175.40): V

Ion 254.40 (254.10 to 255.10): V






#72




1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.35 min Scan# 3599
Delta R.T. -0.00 min
Lab File: VN052908.D
Acq: 14 Dec 2018 23:59

Instrument :
MSVOA_N
ClientSampleId :

Tot Ion:152 Resp: 682850
Ion Ratio Lower Upper
152 100
115 55.6 28.3 85.0
150 157.1 0.0 347.8




Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V





#74


N-ethyl acetate
Concen: 2.660 ug/l
RT: 11.88 min Scan# 3142
Delta R.T. -0.19 min
Lab File: VN052908.D
Acq: 14 Dec 2018 23:59

Tgt Ion: 43 Resp: 37227
Ion Ratio Lower Upper
43 100
70 1.9 34.2 51.4#
55 4.9 20.3 30.5#
61 10.2 20.0 30.0#



Abundance Ion 43.00 (42.70 to 43.70): V
Ion 70.00 (69.70 to 70.70): V
Ion 55.00 (54.70 to 55.70): V
Ion 61.00 (60.70 to 61.70): V





#76

1,2,3-Trichloropropane

Concen: 36.671 ug/l

RT: 12.40 min Scan# 3306

Delta R.T. -0.15 min

Lab File: VN052908.D

Acq: 14 Dec 2018 23:59

Instrument :

MSVOA_N


ClientSampleId :

Tot Ion: 75 Resp: 369447

Ion Ratio Lower Upper

75 100

77 1.2 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VN052827.D (-3348) (-)

Ion 77.00 (76.70 to 77.70): VN052827.D (-3348) (-)

12.40

250000

200000



150000

100000

50000

0

Time--> 12.30 12.40 12.50



#80

1,3,5-Trimethylbenzene

Concen: 0.375 ug/l

RT: 12.73 min Scan# 3408

Delta R.T. -0.00 min

Lab File: VN052908.D


Acq: 14 Dec 2018 23:59

Tgt Ion: 105 Resp: 15434

Ion Ratio Lower Upper

105 100

120 51.0 24.4 73.2



Abundance Ion 105.00 (104.70 to 105.70): VN052827.D (-3391) (-)

Ion 120.00 (119.70 to 120.70): VN052827.D (-3391) (-)

12.73

10000

8000


6000


4000

2000

0

Time--> 12.70 12.75





#81

trans-1,4-Dichloro-2-butene

Concen: 117.805 ug/l

RT: 12.40 min Scan# 3306

Delta R.T. 0.10 min

Lab File: VN052908.D

Acq: 14 Dec 2018 23:59

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion: 75 Resp: 369447

Ion Ratio Lower Upper

75 100

53 0.0 74.7 112.1#




89 0.0 36.1 54.1#



Abundance Ion 74.90 (74.60 to 75.60): VN052908.D (-3306) (-)

Ion 53.00 (52.70 to 53.70): VN052908.D (-3306) (-)

Ion 88.90 (88.60 to 89.60): VN052908.D (-3306) (-)



#84

1,2,4-Trimethylbenzene

Concen: 0.663 ug/l

RT: 13.04 min Scan# 3505

Delta R.T. 0.00 min

Lab File: VN052908.D


Acq: 14 Dec 2018 23:59

Tgt Ion: 105 Resp: 27640

Ion Ratio Lower Upper



105 100


120 45.4 22.9 68.8



Abundance Ion 105.00 (104.70 to 105.70): VN052908.D (-3504) (-)

Ion 120.00 (119.70 to 120.70): VN052908.D (-3504) (-)





#95

Naphthalene

Concen: 1.543 ug/l

RT: 15.13 min Scan# 4155

Delta R.T. -0.00 min


Lab File: VN052908.D

Acq: 14 Dec 2018 23:59

Instrument :

MSVOA_N

ClientSampleId :




Tot Ion:128 Resp: 38520

Ion Ratio Lower Upper

128 100

127 12.3 10.2 15.4

129 11.7 8.6 13.0



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

