

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121418\
 Data File : VN052883.D
 Acq On : 14 Dec 2018 12:35
 Operator : MD\SY
 Sample : J6401-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 002-1212A-SEWER-OUTFALL-002

Manual Integrations
 APPROVED

MMDadoda
 12/17/2018 3:05:36 PM

Quant Time: Dec 17 15:17:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N112618W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Nov 27 00:48:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.79	128	133238m	30.00	ug/l	0.59
28) 1,4-Difluorobenzene	8.91	114	722739	30.00	ug/l	0.32
57) Chlorobenzene-d5	11.48	117	584350	30.00	ug/l	0.07

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.45	65	267722	29.49	ug/l	0.42
Spiked Amount 30.000	Range 50 - 169		Recovery =	98.30%		
60) 4-Bromofluorobenzene	12.44	95	245252	27.45	ug/l	0.04
Spiked Amount 30.000	Range 56 - 143		Recovery =	91.50%		
63) Toluene-d8	10.25	98	799685	29.51	ug/l	0.16
Spiked Amount 30.000	Range 66 - 137		Recovery =	98.37%		

Target Compounds

	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	4.68	58	619957m	884.993	ug/l	
25) Chloroform	7.91	83	56342m	3.742	ug/l	
41) Bromodichloromethane	9.62	83	16055m	1.512	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121418\
Data File : VN052883.D
Acq On : 14 Dec 2018 12:35
Operator : MD\SY
Sample : J6401-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
002-1212A-SEWER-OUTFALL-002

Manual Integrations
APPROVED

MMDadoda
12/17/2018 3:05:36 PM

Quant Time: Dec 17 15:17:11 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N112618W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Nov 27 00:48:33 2018
Response via : Initial Calibration

