

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121418\
 Data File : VN052895.D
 Acq On : 14 Dec 2018 18:12
 Operator : MD\SY
 Sample : J6403-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1-A

Quant Time: Dec 15 03:38:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 13 01:08:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1324259	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2026210	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1827587	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	741770	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	665800	47.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.56%	
35) Dibromofluoromethane	7.59	113	517370	45.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.48%	
50) Toluene-d8	10.09	98	2493318	50.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.68%	
62) 4-Bromofluorobenzene	12.40	95	823617	49.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.14%	
Target Compounds						
16) Acetone	3.81	43	21674	6.854	ug/l	97
20) Methylene Chloride	4.56	84	515047	36.190	ug/l	97
43) Isopropyl Acetate	8.17	43	77114	3.355	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121418\
Data File : VN052895.D
Acq On : 14 Dec 2018 18:12
Operator : MD\SY
Sample : J6403-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
TP-1-A

Quant Time: Dec 15 03:38:24 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 13 01:08:35 2018
Response via : Initial Calibration

