

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121418\
 Data File : VN052897.D
 Acq On : 14 Dec 2018 19:05
 Operator : MD\SY
 Sample : J6403-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2-A

Quant Time: Dec 15 03:41:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 13 01:08:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1405381	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2159467	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1832620	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	696646	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	725958	48.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.16%	
35) Dibromofluoromethane	7.59	113	559771	46.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.88%	
50) Toluene-d8	10.09	98	2584046	49.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.88%	
62) 4-Bromofluorobenzene	12.40	95	799277	45.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.26%	
Target Compounds						
16) Acetone	3.82	43	26228	7.815	ug/l	89
20) Methylene Chloride	4.56	84	5371881	355.673	ug/l	99
43) Isopropyl Acetate	8.17	43	80258	3.277	ug/l #	80
95) Naphthalene	15.13	128	76164	2.991	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121418\
Data File : VN052897.D
Acq On : 14 Dec 2018 19:05
Operator : MD\SY
Sample : J6403-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
TP-2-A

Quant Time: Dec 15 03:41:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 13 01:08:35 2018
Response via : Initial Calibration

