

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121418\
 Data File : VN052898.D
 Acq On : 14 Dec 2018 19:32
 Operator : MD\SY
 Sample : J6403-16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2-B

Quant Time: Dec 15 03:43:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 13 01:08:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1375370	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2146172	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1907443	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	783228	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	711185	48.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.26%	
35) Dibromofluoromethane	7.59	113	542084	45.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.50%	
50) Toluene-d8	10.09	98	2605351	50.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.32%	
62) 4-Bromofluorobenzene	12.40	95	854324	48.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.08%	
Target Compounds						
16) Acetone	3.82	43	33765	10.280	ug/l	98
20) Methylene Chloride	4.56	84	47584	3.219	ug/l	97
25) 2-Butanone	6.84	43	11214	2.200	ug/l	100
43) Isopropyl Acetate	8.17	43	78986	3.245	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121418\
Data File : VN052898.D
Acq On : 14 Dec 2018 19:32
Operator : MD\SY
Sample : J6403-16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
TP-2-B

Quant Time: Dec 15 03:43:47 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 13 01:08:35 2018
Response via : Initial Calibration

