

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121418\
 Data File : VN052903.D
 Acq On : 14 Dec 2018 21:45
 Operator : MD\SY
 Sample : J6403-36
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 28 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-5-A

Quant Time: Dec 15 03:54:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 13 01:08:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1403433	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2157665	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1921613	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	795836	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	711801	47.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.40%	
35) Dibromofluoromethane	7.59	113	564153	46.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.68%	
50) Toluene-d8	10.09	98	2628511	50.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.68%	
62) 4-Bromofluorobenzene	12.40	95	869482	49.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.28%	
Target Compounds						
16) Acetone	3.82	43	34776	10.376	ug/l	96
20) Methylene Chloride	4.56	84	50437	3.344	ug/l	96
25) 2-Butanone	6.84	43	8072	1.552	ug/l	93
43) Isopropyl Acetate	8.17	43	74757	3.055	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121418\
Data File : VN052903.D
Acq On : 14 Dec 2018 21:45
Operator : MD\SY
Sample : J6403-36
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 28 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
TP-5-A

Quant Time: Dec 15 03:54:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 13 01:08:35 2018
Response via : Initial Calibration

