

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121418\
 Data File : VN052909.D
 Acq On : 15 Dec 2018 00:26
 Operator : MD\SY
 Sample : J6403-60
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 34 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-8-A

Quant Time: Dec 15 04:04:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 13 01:08:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1410337	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2172669	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1879629	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	730593	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	712044	47.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.96%	
35) Dibromofluoromethane	7.59	113	578721	47.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.44%	
50) Toluene-d8	10.09	98	2608580	49.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.22%	
62) 4-Bromofluorobenzene	12.40	95	818814	45.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.90%	
Target Compounds						
16) Acetone	3.82	43	42764	12.697	ug/l	95
20) Methylene Chloride	4.56	84	168036	11.087	ug/l	98
43) Isopropyl Acetate	8.17	43	70760	2.871	ug/l #	77
95) Naphthalene	15.13	128	91837	3.439	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121418\
Data File : VN052909.D
Acq On : 15 Dec 2018 00:26
Operator : MD\SY
Sample : J6403-60
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 34 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
TP-8-A

Quant Time: Dec 15 04:04:49 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 13 01:08:35 2018
Response via : Initial Calibration

