

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121418\
 Data File : VN052911.D
 Acq On : 15 Dec 2018 1:19
 Operator : MD\SY
 Sample : J6390-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 36 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 30081

Quant Time: Dec 15 04:21:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 13 01:08:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1347783	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2071240	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1814823	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	756513	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	679902	47.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.88%	
35) Dibromofluoromethane	7.59	113	540688	46.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.54%	
50) Toluene-d8	10.09	98	2506207	50.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.00%	
62) 4-Bromofluorobenzene	12.40	95	818840	48.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.42%	
Target Compounds						
8) Diethyl Ether	3.41	74	58428	7.755	ug/l	100
11) Tert butyl alcohol	4.77	59	10428	13.023	ug/l #	81
16) Acetone	3.82	43	37228	11.567	ug/l	98
20) Methylene Chloride	4.56	84	53941	3.724	ug/l	99
52) Toluene	10.16	92	73300	2.114	ug/l	97
68) m/p-Xylenes	11.62	106	29447	1.213	ug/l	96
84) 1,2,4-Trimethylbenzene	13.04	105	54635	1.184	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121418\
Data File : VN052911.D
Acq On : 15 Dec 2018 1:19
Operator : MD\SY
Sample : J6390-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 36 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
30081

Quant Time: Dec 15 04:21:34 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 13 01:08:35 2018
Response via : Initial Calibration

