

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121418\
 Data File : VN052912.D
 Acq On : 15 Dec 2018 1:46
 Operator : MD\SY
 Sample : J6390-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 37 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 30082

Quant Time: Dec 15 04:24:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 13 01:08:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1344831	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2079973	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1817660	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	757060	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	672891	47.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.12%	
35) Dibromofluoromethane	7.59	113	536523	46.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.42%	
50) Toluene-d8	10.09	98	2514971	49.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.92%	
62) 4-Bromofluorobenzene	12.40	95	823320	48.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.54%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	4.77	59	23900	29.913	ug/l #	80
16) Acetone	3.82	43	91437	28.471	ug/l	98
18) Methyl Acetate	4.33	43	71348	6.053	ug/l	99
20) Methylene Chloride	4.55	84	32697	2.262	ug/l	98
25) 2-Butanone	6.84	43	57657	11.570	ug/l	98
40) Benzene	8.04	78	1014856	17.653	ug/l	99
52) Toluene	10.16	92	390515	11.216	ug/l	99
68) m/p-Xylenes	11.62	106	70218	2.888	ug/l	99
69) o-Xylene	11.95	106	24264	1.022	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121418\
Data File : VN052912.D
Acq On : 15 Dec 2018 1:46
Operator : MD\SY
Sample : J6390-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 37 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
30082

Quant Time: Dec 15 04:24:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 13 01:08:35 2018
Response via : Initial Calibration

